

JUN 5 1981

FCMLB:MAV  
030-01247  
(07412)

Bridgeport Hospital  
ATTN: James E. Bond, Ph.D.  
Radiation Physics Department  
267 Grant Street  
Bridgeport, Connecticut 06602

Gentlemen:

Enclosed is Amendment No. 40 to Byproduct Material License No. 06-01060-01, authorizing the requested changes submitted in your letter of March 24, 1981.

Please note that your license incorporates the use of NRC's model ALARA program contained in Appendix O of Regulatory Guide 10.8 (Rev. 1), copy enclosed. This action was taken because your request did not contain information concerning the ALARA program that you will follow.

As of August 15, 1980, our licensees were requested to submit an ALARA program with any significant amendment, renewal or new byproduct material license application. If you wish to develop and submit an alternative ALARA program that is at least equivalent to the model ALARA program, your license may be amended accordingly.

Since your license is issued in the name of Bridgeport Hospital, your amendment request should have been signed by the hospital administrator. In the future, please be certain that all correspondence, requests for amendment or renewal, etc., are signed by the hospital administrator.

We wish you continued success with your nuclear medicine service.

Sincerely,

John E. Glenn  
Material Licensing Branch  
Division of Fuel Cycle and  
Material Safety

Enclosure: As Stated

8512200300 850724  
REG1 LIC30  
06-01060-01 PDR

OFFICE	FCMLB						
SURNAME	GLENN						
DATE	6/4/81						

OFFICIAL RECORD COPY