

NOTICE OF VIOLATION

Cleveland Clinic Foundation

License No. 34-00466-01

As a result of the inspection conducted on December 5 and 6, 1985, and in accordance with the "General Policy and Procedures for NRC Enforcement Actions," 10 CFR Part 2, Appendix C (1985), the following violation was identified:

License Condition No. 23 requires that licensed material be possessed and used in accordance with statements, representation, and procedures contained in application dated July 23, 1979 and referenced letters.


This application states that GM radiation surveys will be performed daily in dose preparation and patient injection areas and that radiation areas will be wipe tested.

Contrary to the above, from statements of licensee representatives and the absence of records, GM area surveys have not been performed daily or wipe testing conducted weekly in either of your hot labs (Nuclear Medicine and Cardiac). Specifically, no records of daily surveys were available for periods up to ten months, and that the requirements for daily area contamination surveys has not been met in both Nuclear Medicine and Cardiac Departments.

Pursuant to the provisions of 10 CFR 2.201, you are required to submit to this office within thirty days of the date of this Notice a written statement or explanation in reply, including for each violation: (1) corrective action taken and the results achieved; (2) corrective action to be taken to avoid further violations; and (3) the date when full compliance will be achieved. Consideration may be given to extending your response time for good cause shown.

DEC 8

Dated _____


D. G. Wiedeman, Chief
Nuclear Materials Safety
Section 1

8512240058 851220
REG3 LIC30
34-00466-01 PDR