

APPENDIX A

NOTICE OF VIOLATION

Multi Circuits, Inc.
Manchester, Connecticut 06040

Docket No. 030-20677
License No. 06-20653-01

As a result of the inspection conducted on August 16, 1985, and in accordance with the NRC Enforcement Policy (10 CFR 2, Appendix C), the following violations were identified:

- A. Condition 16 of License No. 06-20653-01 requires that licensed material be possessed and used in accordance with the statements, representations and procedures contained in an application dated November 4, 1983.

Items 13.b of this application, at end of second paragraph, requires that the Lixiscope be secured at all times to prevent any unauthorized use.

Contrary to the above, on August 16, 1985, the date of the inspection, the lixiscope was located in an open area, accessible to employees in the licensee's facility. The key for unlocking the lixiscope, enabling it to be used, was in the lock.

This is a Severity Level IV violation. (Supplement IV)

- B. Condition 12 of License No. 06-20653-01 requires that licensed material be used by, or under the supervision and in the physical presence of, two named individuals.

Contrary to the above, as of August 16, 1985, the Lixiscope containing iodine-125, was being routinely used by two drill operators, and occasionally used by another employee who were not named in the license.

This is a Severity Level IV violation. (Supplement VI)

Pursuant to the provisions of 10 CFR 2.201, Multi Circuits, Inc. is hereby required to submit to this office within thirty days of the date of the letter which transmitted this Notice, a written statement or explanation in reply, including: (1) the corrective steps which have been taken and the results achieved; (2) corrective steps which will be taken to avoid further violations; and (3) the date when full compliance will be achieved. Where good cause is shown, consideration will be given to extending this response time.

8512260021 851203
REG1 LIC30
06-20653-01 PDR