

File
THRU :Glen D. Brown, Senior Radiation Specialist
Region IV, Division of Compliance, Denver
Herman J. Paas, Jr., Radiation Specialist
Region IV, Division of Compliance, Denver

ORIGINAL SIGNED BY
G. D. BROWN

February 27, 1968

Original Signed By
Herman J. Paas

DEPT. OF THE ARMY, FITZSIMONS GENERAL HOSPITAL AND U. S. ARMY
MEDICAL RESEARCH AND NUTRITION LABORATORY, DENVER, COLORADO
LICENSE NOS. 05-00046-13 AND -14 - HEALTH AND SAFETY EVALUATION
INSPECTION CONDUCTED FEBRUARY 15, 16, and 19, 1968

The subject licensee uses byproduct material at the U. S. Army Medical Research and Nutrition Laboratory under broad license authorization for In Vitro and research studies related to lower animals, and uses byproduct material under specific medical authorizations in the radioisotope clinic which is operated in conjunction with Fitzsimons General Hospital. Inspection of the licensee's program revealed three items of noncompliance. Two of these items were related to License No. 05-00046-13 and involved overpossession and unauthorized use. Neither of these items of noncompliance offered any apparent threat to the health and safety of the public or the licensee's employees as the amounts involved and the uses of the material represent generally accepted practices and, in this particular case, the items of noncompliance result from the licensee's failure to process the necessary amendments to the existing license. The final item of noncompliance involved failure of the licensee to record radiation surveys that were performed in respect to the Co-60 irradiator licensed under License No. 05-00046-14. A survey performed by the inspector at the time of the inspection confirmed values that were stated verbally by the licensee and the results of this survey define that there was no apparent threat to the health and safety of the public or the licensee's employees during the operation of this unit.

Based on the findings and observations made during this inspection, it appears that the licensee's program is administered in a manner such that there is no apparent threat to the health and safety of the public or the licensee's employees.

cc: J. R. Roeder, CO:HQ ✓