

file

July 29, 1985

Frederick C. Smith Medical Center
Attn: Mr. Robert R. Tracht
Administrator
1050 Delaware Ave.
Marion, OH 43302

This is in reference to a report dated July 19, 1985, signed by Robert R. Tracht, describing the results of a radiation survey of your cobalt-60 teletherapy unit. In order to complete our evaluation of the report, we need the following additional information:

1. We note that the name listed on the letterhead of the report differs from the name under which your license is issued. Please specify the correct name to be used under Item 1 (name) of your license.


2. Please resubmit your facility diagram keyed to the survey results. The diagram currently referenced in your license is dated September 17, 1981 and several survey points listed in your report do not clearly correlate with this diagram.

3. Please clarify what standards (i.e. radionuclide, activity and source accuracy) and procedures were used to calibrate your survey instrument referred to in this report. If your standards and procedures remain the same as those listed in your March 9, 1982 letter to the NRC, you may simply so state.

Please submit the above information in duplicate within thirty (30) days so that we may complete our evaluation of your report. Your response should reference Control No. 79375.

Sincerely,

8511180082 850924
REG3 LIC30
34-00179-03 PDR


William J. Adam, Ph.D.
Material Licensing Section

REGION 3
WJAdam

OFFICE	WJA	07 / 85				
SURNAME						
DATE						