

## MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

## Licensee

1. Canton Analytical Laboratory, Inc.

3. License number 21-19295-02

2. 153 Elder Street  
P.O. Box 1129  
Ypsilanti, MI 48197

4. Expiration date September 30, 1990

5. Docket or  
Reference No. 030-286346. Byproduct, source, and/or  
special nuclear material7. Chemical and/or physical  
form8. Maximum amount that licensee  
may possess at any one time  
under this license

A. Nickel-63

A. Foil source  
(Shimadzu Scientific  
Instruments, Inc. detector  
cell Model No. EDC-7)A. No single cell  
to exceed 15  
millicuries each

B. Nickel-63

B. Plated source  
(Hewlett-Packard  
Model 18803-60520 detector  
cell)B. No single cell  
to exceed 15  
millicuries

## 9. Authorized Use

- A. To be used in a Shimadzu Scientific Instruments gas chromatograph(s) for sample analysis.
- B. To be used in a Hewlett-Packard gas chromatograph for sample analysis.

CONDITIONS

10. Licensed material may be used at Canton Analytical Laboratory, 153 Elder Street, Ypsilanti, Michigan and at temporary job sites of the licensee anywhere in the United States where the U. S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."
12. Licensed material shall be used by, or under the supervision of, Moreton B. Price or Peter W. Rekshan.

8511180087 850924  
REG3 LIC30 PDR  
21-19295-02

COPY 2

**MATERIALS LICENSE**  
SUPPLEMENTARY SHEET

License number

12-19295-02

Docket or Reference number

030-28634

13. In lieu of using the conventional radiation caution colors (magenta or purple on yellow background) as provided in Section 20.203(a)(1), Title 10, Code of Federal Regulations, Part 20, the licensee is hereby authorized to label detector cells and cell baths, containing licensed material and used in gas chromatography devices, with conspicuously etched or stamped radiation caution symbols without a color requirement.
14. Detector cells containing licensed material shall not be opened or the foil sources removed from the detector cell by the licensee.
15. Maintenance, repair, cleaning, replacement and disposal of foils contained in detector cells shall be performed only by the device manufacturer or other persons specifically authorized by the Commission or an Agreement State to perform such services.
16. Except as otherwise specified in this license, the licensee shall have available and follow the instructions contained in the manufacturer's instruction manual for the chromatography device.
17. The licensee shall conduct a physical inventory every six (6) months to account for all cells received and possessed under the license. The records of the inventories shall be maintained for two (2) years from the date of the inventory for inspection by the Commission, and shall include the quantities and kinds of byproduct material, manufacturer's name and model numbers, location of cells and the date of the inventory.
18. The licensee shall dispose of sources by returning to the manufacturer when their use has been discontinued.
19.
  - A. Each chromatograph detector containing Nickel 63 shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a detector received from another person shall not be put into use until tested.
  - B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the surfaces of the device in which the foil is mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
  - C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the foil from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the test with the U. S. Nuclear Regulatory Commission, Region III, 799 Roosevelt Road, Glen Ellyn, Illinois 60137, describing the equipment involved, the test results, and the corrective action taken.

**MATERIALS LICENSE**  
SUPPLEMENTARY SHEET

License number

12-19295-02

Docket or Reference number

030-28634

- D. Tests for leakage and/or contamination shall be performed by the licensee using an NRC approved leak test kit for analysis by persons specifically authorized by the Commission or an Agreement State to perform such services or by other persons specifically authorized by the Commission or an Agreement State to perform such services.
20. The licensee may transport licensed material or deliver licensed material to a carrier for transport in accordance with the provisions of Title 10, Code of Federal Regulations, Part 71, "Packaging of Radioactive Material for Transport and Transportation of Radioactive Material Under Certain Conditions."
21. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated October 25, 1979; letter received March 14, 1980; and letters dated June 12, 1981, December 18, 1981, February 25, 1982 and April 10, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.



For the U.S. Nuclear Regulatory Commission

SEP 24 1985

Date \_\_\_\_\_

Original Signed  
By George M. McCann  
Materials Licensing Section, Region III

COPY