

SEP 10 1985

Foth & Van Dyke and Associates
ATTN: Neil A. Van Dyke
Division General Manager
2737 South Ridge Road
P.O. Box 19012
Green Bay, WI 54307-9012

Re: Application Dated June 20, 1985 For An NCR License

Gentlemen:

In order to complete your licensing request it will be necessary for you to submit the following information:

1. Your application indicates that you wish to possess (2) Perkin-Elmer detector cells (Model No. 0331-3368) for use in a gas chromatograph. Be advised that our Sealed Source and Device Registry does not contain the Perkin-Elmer detector cell model number specified in your application. Please re-check your records and verify the detector cell model number you wish to possess.
2. Describe your procedures for:
 - a. Disposal of the foil when its use has been discontinued. If this disposal is by return to the manufacturer, please so state.
 - b. Cleaning detector cells and/or removal and exchange of detector cell foil. If you follow the manufacturer's recommended procedures, please provide a copy of the procedures. If detector cells are returned to your supplier for cleaning of detector cells and/or removal and exchange of foil, please specify this.
 - c. Leak testing the nickel-63 foil. If you will use a commercial leak test kit, please provide the name of the supplier and model number of the kit. If you will conduct your own leak tests, please provide descriptions of the following:
 - (1) The materials and procedures used for collecting leak test samples.
 - (2) The name of the manufacturer and model number of the measuring instrument used to analyze leak test samples.

8511180299 851113
REG3 LIC30 PDR
48-18916-02

- (3) Your procedures for calibration of the measuring instrument including a sample calculation showing how leak test results are converted to microcuries.

If you have any questions or require clarification on any of the information stated above, you may contact us at (312) 790-5625.

We will continue our review of your application upon receipt of this information. Please reply in duplicate, within 30 days, and refer to Control Number 78244.

Sincerely,

Original Signed By
Patricia J. Whiston
Materials Licensing Section

RIID
Whiston/cm
09/09/85