

NOV 12 1985

Docket No. 030-19183

License No. 06-19749-01

Diagnostic Radiology Associates
ATTN: Dr. Berg
134 Grandview Avenue
Waterbury, Connecticut 06708

Gentlemen:

Subject: Inspection No. 030-19183/85-01

This refers to the routine safety inspection conducted by Judith Joustra and John Jensen of this office on October 8, 1985 of activities authorized by NRC License No. 06-19749-01 and to the discussions of our findings held by Ms. Joustra with yourself at the conclusion of the inspection.

The inspection was an examination of activities conducted under your license as they relate to radiation safety and to compliance with the Commission's rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, measurements made by the inspectors, and observations by the inspectors.

Based on the results of this inspection, it appears that certain of your activities were not conducted in full compliance with NRC requirements, as set forth in the Notice of Violation, enclosed herewith as Appendix A. These violations have been categorized by severity level in accordance with the revised NRC Enforcement Policy (10 CFR 2, Appendix C) published in the Federal Register Notice (49 FR 8583) dated March 8, 1984. You are required to respond to this letter and in preparing your response, you should follow the instructions in Appendix A.

From the discussion of our findings at the conclusion of the inspection we understand that area survey procedures will include a daily survey with an appropriate low-range survey meter and a weekly series of wipe tests. Please confirm our understanding in your reply to this letter.

In accordance with Section 2.790 of the NRC's "Rules of Practice," Part 2, Title 10, Code of Federal Regulations, a copy of this letter and your reply will be placed in the Public Document Room.

8511180433 851112
REG1 LIC30 PDR
06-19749-01

OFFICIAL RECORD COPY

DL DIAG RAD - 0001.0.0
11/06/85

RETURN ORIGINAL TO
REGION I

11
IE:07

NOV 12 1985

Diagnostic Radiology Associates

2

The responses directed by this letter and the accompanying Notice are not subject to the clearance procedures of the Office of Management and Budget as required by the Paperwork Reduction Act of 1980, PL 96-511.

Your cooperation with us in this matter is appreciated.

Sincerely,

Original Signed By:

John E. Glenn

John E. Glenn, Ph.D, Chief
Nuclear Materials Safety Section B,
Division of Radiation Safety
and Safeguards

Enclosure: Appendix A, Notice of Violation

cc w/encl:

Public Document Room (PDR)

Nuclear Safety Information Center (NSIC)

State of Connecticut

bcc w/encl:

Region I Docket Room (w/concurrences)

Management Assistant, DRMA

RI:DRSS
Joustra/ca

RI:DRSS
Glenn

11/8/85

11/8/85

OFFICIAL RECORD COPY

DL DIAG RAD - 0002.0.0
11/06/85