

OCT 10 1985

Lake-Cook Orthopedic Associates  
ATTN: Michael P. Young, M.D., S.C.  
President  
450 West Highway 22  
Barrington, IL 60010

Gentlemen:

Enclosed is Amendment No. 01 to your NRC License No. 12-24554-01 in accordance with your request.

Please note that the Amersham Corporation Model IMC.P2 has been approved for use with the Nuclear Data, Inc. Model ND1100 Bone Densitometer since September 25, 1985; therefore it has been added to Subitem 7.A. of your license as requested. However, according to the sealed source and device registry safety evaluation, the ND1100 is approved for 200 millicuries of iodine-125, consequently, Subitem 8.A. of your license has not been amended to increase your possession limit to 250 millicuries.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By  
J.R. Madera  
Materials Licensing Section

Enclosure(s): Amendment No. 01

8511180443 851010  
REQ3 LIC30  
12-24554-01 PDR

RITI

Madera/cm

10/07/85

*AM m*  
*10/07/85*