

NMSS:JEW  
Control No. 60304


JUN 13 1985

Missoula Community Hospital  
ATTN: Grant M. Winn  
Executive Director  
2827 Fort Missoula Road  
Missoula, Montana 59801

Gentlemen:

This is in reference to your request for a byproduct material license renewal in the name of Missoula Community Hospital. In order for us to complete our review of your request, please supply the following:

- dl* 1. 10 CFR Part 35.14(b)(1) requires in part that Group-licensed radiopharmaceuticals administered to humans must be labeled, packaged, and distributed by individual licensees holding a specific license from NRC or an Agreement State authorizing distribution. Receipt of Group-licensed materials from individuals not possessing a distribution license specified above must be approved by NRC on a case-by-case basis. You should request authorization to receive group radiopharmaceuticals from Saint Patrick Hospital (License No. 25-16773-02).
- dl* 2. Radiation workers (technologists, etc.) must receive instruction as specified in 10 CFR 19.12 (enclosed). Note that many of these items pertain to circumstances at your particular institution; therefore, you may not assume that this instruction has been adequately covered by prior occupational training, Board certification, etc. Outline and submit your program for providing the necessary instruction initially when workers are hired and annually thereafter on a refresher basis.
- dl* 3. Ancillary personnel (clerical, nursing, housekeeping, security, etc.) whose duties may require them to work in the vicinity of radioactive material (whether escorted or not) need to be informed about radiation hazards and appropriate precautions. Outline your method to assure that these employees receive the necessary instruction initially when they are hired and annually thereafter on a refresher basis.

  
NMSS  
JEWWhitten;df  
6/4/85

C:NMSS  
RJEverett  
6/7/85

8511180609 850910  
REG4 LIC30  
25-18361-01 PDR

4. You should develop written instructions for the receipt of shipments of radioactive materials. Your instructions should be adequate to assure that licensed material is secured at all times against unauthorized removal, and that radiation levels in unrestricted areas do not exceed the limits specified in 10 CFR 20.105. In addition, these instructions should explain what to do if a package is found to be damaged or leaking, including the names and telephone numbers of responsible persons to be contacted. Sample instructions NRC finds acceptable are outlined in Appendix E of Regulatory Guide 10.8. Submit your written procedures for NRC staff review and approval.

We will continue our review of your application upon receipt of this information. Please reply in duplicate and refer to Control No. 60304.

Sincerely,  
*Original Signed By.*  
Jack E. Whitten

Jack E. Whitten  
Nuclear Materials Safety Section

bcc w/o enclosures:  
RLBangart, D/DRSP