

OCT 8 1985

The Christ Hosptial  
ATTN: Jack M. Cook  
President  
2139 Auburn Avenue  
Cincinnati, OH 45219

Gentlemen:

Enclosed is Amendment No. 35 renewing your NRC License No. 34-03831-02 in accordance with your request.

Several new conditions have been incorporated in your license. In view of this, please review your license carefully to assure that you understand and are in compliance with the terms and conditions contained therein.

Please note that you are subject to the Confirmatory Order dated September 11, 1985, modifying your license to incorporate the radioactive material implant procedures outlined in your letter dated June 28, 1985. This Order should be adhered to until such time that the order is modified or rescinded. Amendment No. 35 does not refer to the above mentioned implant procedures; they should be incorporated in your license through a formal amendment request thereby eliminating the need for the Confirmatory Order.

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.

8511180648 851008  
REG3 LIC30  
34-03831-02 PDR

5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.
6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.
7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By  
J.R. Madera  
Materials Licensing Section

Enclosure(s):

1. Amendment No. 35
2. "Guidelines For Bioassay  
Requirements For Tritium",  
October 19, 1977

RH  
JRM  
Madera/cm  
10/01/85

## CONVERSATION RECORD

TIME

11:30 a

DATE

5/22/85

TYPE

☐ VISIT☐ CONFERENCE☒ TELEPHONE☐ INCOMING☒ OUTGOING

ROUTING

NAME/SYMBOL

INT

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT  
WITH YOU

Clifford H. Born, M.S., RSO

ORGANIZATION (Office, dept., bureau,  
etc.)

Christ Hospital

TELEPHONE NO.

513/369-2323

SUBJECT

Telephone deficiency - Renewal application

SUMMARY

- ① Please refer to App. B of Reg. Guide 10.8 and submit procedures to meet quarterly (Med. Isotopes Com.)
- ② Please refer to Appendix F of Reg. Guide 10.8 and submit procedures for measuring exposures of packages received (Rad. mat.) at 3 feet.
- ③ Please submit your procedures to perform Bioassays in accordance to Reg. Guide 10.8. and Handling procedures.
- ④ Please submit your procedures for handling brachytherapy sealed sources - Procedures equivalent to Appendix L of Guide 10.8.
- ⑤ Please submit your ALARA Program in accordance with Reg Guide 10.8 appendix O.
- ⑥ Please submit procedures for Rad. material use in animals (facilities; instruction to caretakers; procedures for handling animals, waste & cages; security).

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

ACTION TAKEN

SIGNATURE

TITLE

DATE