

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s); and to import such byproduct and source material. This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

"OFFICIAL RECORD COPY"

Licensee

1. Canfield Radiology Associates
Canfield Office Condominium

3. License number 29-20881-01

2. 586 Pompton Avenue
Cedar Grove, New Jersey 07068

4. Expiration date September 30, 1990

5. Docket or
Reference No. 030-287926. Byproduct, source, and/or
special nuclear material7. Chemical and/or physical
form8. Maximum amount that licensee
may possess at any one time
under this license

A. Gadolinium 153

A. Sealed source (Lunar
GD series)A. Two sources not to exceed
1,500 millicuries each
source

9. Authorized use

A. For use in Lunar Radiation Corporation DP3 bone mineral analyzer.

CONDITIONS

10. Licensed material shall be used only at the licensee's facilities, Canfield Office Condominium, 586 Pompton Avenue, Cedar Grove, New Jersey.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions, and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."
12. Licensed material listed in Item 6 above is authorized for use by the following individual(s) for the materials and uses indicated:

Patrick J. Conte, M. D.

Gadolinium 153 sealed sources for diagnosing
bone maladies

Andrew J. DeRogatis, M.D.

Gadolinium 153 sealed sources for diagnosing
bone maladies8511180695 850916
REG1 LIC30
29-20881-01 PDR

ML10

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

29-20881-01

Docket or Reference number

030-28792

(continued)

CONDITIONS

13. A. (1) Each sealed source containing licensed material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed one year. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a sealed source received from another person shall not be put into use until tested.
- (2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the test with the U. S. Nuclear Regulatory Commission, Region I, 631 Park Avenue, King of Prussia, Pennsylvania 19406, describing the equipment involved, the test results, and the corrective action taken.
- D. The licensee is authorized to collect leak test samples in accordance with instructions provided by Lunar Radiation Corporation for analysis by Siemens, or Teledyne, Incorporated. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Commission or an Agreement State to perform such services.
14. Sealed sources containing licensed material shall not be opened or removed from their respective source holders by the licensee.
15. Maintenance, repair, installation, replacement, and disposal of sealed sources containing licensed material used in devices shall be performed only by Lunar Radiation Corporation or by other persons especially authorized by the Commission or an agreement State to perform such services. Alternatively, source exchanges may be performed by authorized users who have received actual demonstrations and instructions on installing and replacing sources. The licensee shall maintain for inspection by the Nuclear Regulatory Commission records of such training.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number

29-20881-01

Docket or Reference number

030-28792

(continued)

CONDITIONS

16. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated June 27, 1985, and letter dated August 26, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.



For the U.S. Nuclear Regulatory Commission

Original Signed By:

John D. Kinneman

By

Nuclear Materials Safety and
Safeguards Branch, Region I
King of Prussia, Pennsylvania 19406

Date SEP 16 1985