

HOWARD-DONLEY ASSOCIATES, INC.  
SOIL AND FOUNDATION ENGINEERING  
P.O. BOX 3679, CASPER, WYOMING, 82602  
(307) 235-0551



August 27, 1985


USNRC  
Region IV  
Material Safety Section  
Parkway Central Plaza Building  
611 Ryan Plaza Drive, Suite 1000  
Arlington, Texas 76011

RE: Control Number 460702  
Additional Information  
Byproduct Material License

Gentlemen:

Attached is a copy of your August 22 letter requesting additional information concerning the referenced subject. Also attached your check list and our response. Please contact our office if you need any additional information. Your prompt attention is appreciated.

Sincerely,

A handwritten signature in cursive script that reads "Larry G. Clark".

Larry G. Clark  
Radiation Protection Officer

attachments

8511190108 850926  
REG4 LIC30  
49-19334-02 PDR

460702

*Don't but let go*

ADDITIONAL REQUESTED INFORMATION  
HOWARD-DONLEY ASSOCIATES, INC.  
CONTROL NUMBER 450702

Item 9, Page 7

The gauge will be stored in our laboratory located at 240 South Wolcott, Casper, Wyoming. Only authorized laboratory personnel, trained in the use of the gauge, are allowed in the laboratory. The laboratory is locked when no one is present.

✓ Item 10.5, Pages 9 & 10

The gauge will be transported to job sites in work vehicles. Vehicles display radioactive material signs on the outside when they are transporting the gauge. Packaging and transporting the device will be carried out in accordance with applicable DOT regulations.

Item 10.6, Page 10

Personnel authorized to use the gauge have written information concerning the following:

Film badges will be worn when using the gauge,

*Don't* { Step-by-step instructions outlining the proper use and care of the gauge are provided (ie: protection from traffic, checking calibration, public protection, etc.),

The gauge shall be stored only in the laboratory or in a locked vehicle during daytime breaks,

The gauge will be adequately secured to prevent jarring when being transported and located as far away from passengers as possible,

Leak tests will be performed by the Radiation Protection Officer,

During an emergency the person using the gauge will notify the RPO and Sheriff's Department. The operator will remain on site and keep people away from the area. The RPO will notify a representative of the NRC.

Item 11, Pages 10 & 11

✓ The gauge will be transferred only to another firm possessing the proper NRC license.