

January 6, 1984

Mr. Charles Stearns
U. S. Nuclear Regulatory Commission
Division I
631 Park Avenue
King of Prussia, Pennsylvania 19406

Dear Mr. Stearns:

This is to confirm our telephone conversation of January 6, 1984. We requested permission to service one of the Cesium-137 irradiators at Frederick Cancer Research Center in Frederick, Maryland. We have two service engineers with factory training on the J. L. Shepherd irradiators. Enclosed is a copy of the certifying letter from the manufacturer of the unit.

We acknowledge your verbal permission to service the unit and appreciate your cooperation in this matter.

Sincerely,

Therapy Services Incorporated

Jerry L. Jogle
President

8511190373 851113
REG1 LIC30 PDR

THERAPY SERVICES INCORPORATED

Service Record

Customer

Name of Contact

FREDERICK CANCER RESEARCH CENTER MR. GREG CLARKE

N.C.I. Telephone 816-95-1436

FREDERICK, M.D. 21701 Date Received JANUARY 6, 1984

Type of Service Requested REPAIR SOURCE DRIVE MECHANISM -
SOURCE WILL NOT GO OFF

Unit Manufacturer J.L. SHEPHERD Model MARK I 68A S/N 1016

Date of Service 1-6-84 Date Completed 1-6-84

Service Performed INSPECTED SOURCE DRIVE MECHANISM -
DRIVE ROD IS DISCONNECTED FROM SOURCE ASSEMBLY -
SOURCE WAS IN FULL "OFF" POSITION, WILL
CONTACT FACTORY REGARDING RECONNECTING SOURCE
ASSEMBLY TO DRIVE ROD - WILL CONTACT GREG
CLARKE WITH FACTORY INFORMATION

Parts Used _____

Recommendations for Future Service _____

Personnel	Dosimeter S/N	Reading
-----------	---------------	---------

<u>L.F. FERRELL</u>	<u>603287</u>	<u>0</u>
---------------------	---------------	----------

<u>J.L. FOGLE</u>	<u>7607D</u>	<u>0</u>
-------------------	--------------	----------

Signed

James Fogle
Therapy Services Incorporated

Date 1-6-84

P.O. No. 121961

SUBMIT SERVICE RECORD TO MR. BUCK CLARKE

BLDG 100

YOUR PERMISSION REQUIRED
FROM NRC - 1-6-84, TO
SERVICE UNIT

TSI Form 102

James Fogle 1-6-84