

From: [Gallagher, Robert](#)
To: [Travis Painter](#); [Loretta Simon](#)
Subject: Williamson Memorial Hospital Termination - Request for Additional Information, MC 618631
Date: Tuesday, May 19, 2020 10:46:00 AM
Attachments: [image003.png](#)

License No. 47-25358-01
Docket No. 03034138
Control No. 618631

Mr. Painter,

I received the letter from Williamson Memorial Hospital yesterday regarding the removal of the sealed sources from the hospital. Included with the letter was a list of all of the sealed sources (nine in total) and the results of the latest leak test performed on the sources. I note that you leak tested four sources, not all nine sources. Please explain why all nine sources were not leak tested prior to shipment, or submit leak test results for the remaining five sources.

Regards,

Robert L. Gallagher
Health Physicist
U.S. NRC, Region I
2100 Renaissance Blvd.
King of Prussia, PA 19406
(610) 337-5182 office
(610) 337-5269 fax

