

U. S. NUCLEAR REGULATORY COMMISSION
MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter 1, Parts 30, 31, 32, 33, 34, 35, 36, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s); and to import such byproduct and source material. This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee		In accordance with application dated May 19, 1980, 3. License number 34-16587-01 is amended in it's entirety to read as follows:
1. City of Massillon Wastewater Treatment Plant		
2. 2695 Erie Street, South Massillon, Ohio 44646		
		4. Expiration date October 31, 1985
		5. Docket or Reference No.
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Cesium-137	A. Sealed source (Ohmart Corporation Model No. A-2102)	A. One source not to exceed 50 millicuries
9. Authorized use		
A. For use in Ohmart Corporation Model SHLM source holder for level measurement.		

CONDITIONS

10. Licensed material shall be used only at City of Massillon, Wastewater Treatment Plant, 2695 Erie Street, South, Massillon, Ohio.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."
12. Licensed material shall be used by, or under the supervision of, Robert A. Bushman or William L. Fichter.

8141280239

MATERIALS LICENSE

Supplementary Sheet

License Number 34-16587-01

Docket or
Reference No. _____

Amendment No. 01

Continued from Page 1

CONDITIONS

13. A(1) Each sealed source containing licensed material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a sealed source received from another person shall not be put into use until tested.
- A(2) The periodic leak test required by this condition does not apply to sealed sources that are stored and not being used. The sources excepted from this test shall be tested for leakage prior to any use or transfer to another person unless they have been leak tested within six months prior to the date of use or transfer.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within five (5) days of the test with the U. S. Nuclear Regulatory Commission, Region III, Office of Inspection and Enforcement, 799 Roosevelt Road, Glen Ellyn, Illinois 60137, describing the equipment involved, the test results, and the corrective action taken.
- D. Tests for leakage and/or contamination shall be performed by the licensee or by other persons specifically authorized by the Commission or an Agreement State to perform such services.
14. Sealed sources containing licensed material shall not be opened or removed from their respective source holders by the licensee.

MATERIALS LICENSE

Supplementary Sheet

License Number 34-16587-01

Docket or
Reference No. _____

Amendment No. 01

Continued from Page 2

CONDITIONS

15. Installation, initial radiation survey of devices, relocation, maintenance, repair, and removal from service of the devices containing licensed material and installation, replacement, and disposal of sealed sources containing licensed material used in the devices shall be performed only by Cimart Corporation or by other persons specifically authorized by the Commission or an Agreement State to perform such services.
16. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6., 7., and 8. of this license in accordance with statements, representations, and procedures contained in application dated May 19, 1980. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

SEP 12 1980

Date _____

For the U. S. Nuclear Regulatory Commission

by Samuel S. Pitts
Material Licensing Branch

Division of Fuel Cycle and
Material Safety
Washington, D.C. 20555

CONTROL NO. 7 9575