



SEWICKLEY VALLEY HOSPITAL, SEWICKLEY, PENNSYLVANIA 15143

John D. Kinneman  
U.S. Nuclear Regulatory Commission  
631 Park Ave.  
King of Prussia, Pennsylvania 19406

NCR License No. 37-11079-01  
Inspection No. 030-03131/85-01

Dear Mr. Kinneman:

Referring to your letter, of November 12, 1985 regarding the above referenced inspection, please note the following:

Item A

Radiation workers were given Inservice on 12/6/84.

Item B

Surveys for licensed material in trash cans are now being made before disposal in the normal trash to ensure that no licensed material will be placed in normal trash.

Item C

One of the survey meters has been fitted with an end-window GM tube and calibrated. The sensitivity has been determined to insure detection of low levels of contamination. (i.e. 200 dpm/100 cm squared)

Item C 2a

The linearity tests have been done according to the procedures submitted in our license applications, (approved), which state that the measured and calculated activities will be plotted versus time.

Item C 2b

The importance of daily consistency checks has been stressed to the Nuclear Medicine Technologists to avoid any recurrence of non-daily checks.

8601230435 860110  
REG1 LIC30  
37-11079-01 PDR

Item C 2c

The net activity of the reference sources will now be recorded to at least two decimal places, where applicable, to insure complete dose calibrator consistency evaluations.

Item C 3

The external surface of the final source containers will now be wiped to check for contamination.

Item C 4a

The importance of daily surveys will be stressed to the Nuclear Medicine Technologists to insure that surveys are being done as required.

Item C 4b

See reference to Item C .

Sincerely,

A handwritten signature in dark ink, appearing to read 'Marvin Wedeen', with a long, sweeping horizontal line extending to the right.

Marvin Wedeen  
Vice President