

Amendment No. 09

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

<p>Licensee</p> <p>1. Uniroyal Tire Company, Inc.</p> <p>2. Uniroyal Road Ardmore, Oklahoma 73401</p>		<p>In accordance with letter dated October 14, 1985</p> <p>3. License number 35-16911-01 is amended in its entirety to read as follows:</p>
		<p>4. Expiration date May 31, 1986</p>
		<p>5. Docket or Reference No. 030-11854</p>
<p>6. Byproduct, source, and/or special nuclear material</p> <p>A. Cesium-137</p> <p>B. Strontium-90</p>	<p>7. Chemical and/or physical form</p> <p>A. Sealed sources</p> <p>B. Sealed sources</p>	<p>8. Maximum amount that licensee may possess at any one time under this license</p> <p>A. See Item 9.A.</p> <p>B. See Item 9.B.</p>

9. Authorized use:

- A. For possession and use in Kay-Ray devices which have been evaluated and approved for licensing purposes and authorized for distribution under a license issued by the Nuclear Regulatory Commission or an Agreement State.
- B. For possession and use in AccuRay devices which have been evaluated and approved for licensing purposes and authorized for distribution under a license issued by the Nuclear Regulatory Commission or an Agreement State.

CONDITIONS

- 10. Licensed material shall be used only at Uniroyal Tire Company, Inc., Uniroyal Road, Ardmore, Oklahoma.

8603240022 851226
REG4 LIC30
35-16911-01 PDR

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number

35-16911-01

Docket or Reference number

030-11854

Amendment No. 09

11. Licensed material shall be used by, or under the supervision of, Ronnie J. Means or Thomas W. Williford.
12.
 - A. Sources contained in devices manufactured by Kay-Ray and AccuRay shall be tested for leakage and/or contamination at intervals not to exceed 6 months. The test may be conducted at 3-year intervals provided the sources have been authorized by the Commission (or an Agreement State) for a 3-year leak test interval. Any source which is received from another person which is not accompanied by a certificate indicating that a test was performed within 6 months before the transfer shall not be put into use until tested.
 - B. Any source in storage and not being used need not be tested. When the source is removed from storage for use or transfer to another person, it shall be tested before use or transfer.
 - C. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the source shall be removed from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the date the leak test result is known with the U.S. Nuclear Regulatory Commission, Region IV, 611 Ryan Plaza Drive, Suite 1000, Arlington, Texas 76011, ATTN: Chief, Radiological and Safeguards Programs Branch. The report shall specify the source involved, the test results, and corrective action taken. Records of leak test results shall be kept in units of microcuries and shall be maintained for inspection by the Commission. Records may be disposed of following Commission inspection.
 - D. The licensee is authorized to collect leak test samples for analysis by Nuclear Sources and Services, Inc., or tests for leakage and/or contamination shall be performed by persons specifically licensed by the Commission or an Agreement State to perform such services.

MATERIALS LICENSE
SUPPLEMENTARY SHEETLicense number
35-16911-01Docket or Reference number
030-11854

Amendment No. 09

13. Sealed sources containing licensed material shall not be opened.
14. Installation, initial radiation survey, relocation, removal from service, maintenance, and repair of devices containing sealed sources shall be performed only by licensee or by persons specifically licensed by the Commission or an Agreement State to perform such services. Installation, replacement, and disposal of sealed sources shall be performed only by persons specifically licensed by the Commission or an Agreement State to perform such services.
15. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 2 years from the date of each inventory.
16. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated January 21, 1980, (received January 27, 1981)
 - B. Letter dated May 7, 1981
 - C. Letter dated October 18, 1982
 - D. Letter dated December 14, 1982
 - E. Letter dated October 14, 1985

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Original Signed By
Jack E. WhittenDate DEC 26 1985

By

Nuclear Materials Safety Section
Region IV
Arlington, Texas 76011