

Nichols Road  
Fitchburg, Massachusetts 01420  
617 343-5000

June 14, 1985

Nuclear Regulatory Commission  
Region I  
631 Park Avenue  
King of Prussia, Pa. 19406

Re: Lic. No. 20-14767-01  
Inspection #84-01 July 9, 1984

030-14395


Dear Sir:

Wipe testing of the sealed cobalt-57 and cesium-137 constancy sources is now implemented. The sources were recently checked and will be tested every six months.

The samples are read in a gamma counter. Appropriate efficiency factors are used. Methodology is sufficiently sensitive enough to detect the presence of .005 uCi removal contamination.

If you have any further questions, do not hesitate to call.

Sincerely,

  
Wesley Munger, Acting Manager

8507120130 850626  
REG1 LIC30  
20-14761-01 PDR

WM:wk