

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number
34-17546-01

Docket or Reference number
030-12943

Amendment No. 04

CORRECTED COPY

S.D. Myers, Inc.
Transformer Consultants Division
P.O. Box 4724
Akron, OH 44310

In accordance with letter dated April 11, 1985, License Number 34-17546-01 is amended as follows:

Items 6., 7., 8., and 9. are amended to add:

6. Byproduct, source,
and/or special nuclear
material

7. Chemical and/or
physical form

8. Maximum amount that
licensee may possess
at any one time
under this license

C. Nickel-63

C. Foil sources
(Shimadzu electron
capture detector Model
Mini-2, containing
New England Nuclear
Model NER-004)

C. No single cell
to exceed 10
millicuries

9. Authorized Use

C. For use in Shimadzu Scientific Instruments gas chromatographs for sample analysis.

Conditions 12., 17. and 22. are amended to read:

12. Licensed material shall be used by, or under the supervision of, William Raber, Joseph J. Kelly or Timothy J. Angel.

17. Exhaust from detector cells containing hydrogen-3 shall be vented through a laboratory fume hood or suitable means designated to reduce potential exposure to personnel to lowest practicable level.

22. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated May 21, 1982; letters dated August 5, 1982 and April 11, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

For the U.S. Nuclear Regulatory Commission

8507120197 850619
REG3 LIC30
34-17546-01 PDR

Date June 19, 1985

Original Signed
By George M. McCann
Materials Licensing Section, Region III

10
MA 30 BK COPY 2
JUN 19 1985