

## MATERIALS LICENSE

Amendment No. 01

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

302598

<p>Licensee</p> <p>1. CGC, Inc.</p> <p>2. 3011 Perry St. Madison, WI 53713</p>	<p>In accordance with the letter dated April 28, 1997</p> <p>3. License Number 48-26614-01 is amended in its entirety as follows:</p> <p>4. Expiration Date December 31, 2004</p> <p>5. Docket or Reference No. 030-33693</p>	
<p>6. Byproduct, Source, and/or Special Nuclear Material</p> <p>A. Cesium-137</p> <p>B. Americium-241</p>	<p>7. Chemical and/or Physical Form</p> <p>A. Sealed source (Troxler Dwg. No. A-102112)</p> <p>B. Sealed source (Troxler Dwg. No. A-102451)</p>	<p>8. Maximum Amount that Licensee May Possess at Any One Time Under This License</p> <p>A. 15 sources not to exceed 10 millicuries each</p> <p>B. 15 sources not to exceed 50 millicuries each</p>

## 9. Authorized Use:

A. and B. To be used in Troxler Model 3400 Series moisture/density gauges.

## CONDITIONS

10. Licensed material may be used and stored at the licensee's facilities located at 3011 Perry Street, Madison, Wisconsin; 205 W. Galena, Suite 100, Milwaukee, Wisconsin; and at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.
11. The Radiation Safety Officer for this license is Kim J. Lewis.
12. Licensed material shall only be used by, or under the supervision and in the physical presence of, either Kim J. Lewis or Donald W. Arenander or individuals who have successfully completed the manufacturer's training program for gauge users, have been instructed in the licensee's routine and emergency operating procedures and who have been designated by the Radiation Safety Officer.

110083  
9706180144 970605  
PDR ADOCK 03033693  
C PDR



COPY

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number

48-26614-01

Docket or Reference Number

030-33693

Amendment No. 01

13. A. Sealed sources and detector cells shall be tested for leakage and/or contamination at intervals not to exceed 6 months or at such other intervals as specified by the certificate of registration referred to in 10 CFR 32.210.
- B. Notwithstanding Paragraph A of this Condition, sealed sources designed to emit alpha particles shall be tested for leakage and/or contamination at intervals not to exceed 3 months.
- C. In the absence of a certificate from a transferor indicating that a leak test has been made within 6 months prior to the transfer, a sealed source or detector cell received from another person shall not be put into use until tested.
- D. Sealed sources need not be leak tested if:
- (i) they contain only hydrogen-3; or
  - (ii) they contain only a radioactive gas; or
  - (iii) the half-life of the isotope is 30 days or less; or
  - (iv) they contain not more than 100 microcuries of beta and/or gamma emitting material or not more than 10 microcuries of alpha emitting material; or
  - (v) they are not designed to emit alpha particles, are in storage, and are not being used. However, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source or detector cell shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- E. The leak test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(b)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. The report shall be filed within 5 days of the date the leak test result is known with the U.S. Nuclear Regulatory Commission, Region III, ATTN: Chief, Nuclear Materials Safety Branch, 801 Warrenville Road, Lisle, Illinois 60532-4351. The report shall specify the source involved, the test results, and corrective action taken.

COPY

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License Number

48-26614-01

Packet or Reference Number

030-33693

Amendment No. 01

- F. The licensee is authorized to collect leak test samples for analysis by Troxler Electronic Laboratories. Alternatively, tests for leakage and/or contamination may be performed by persons specifically licensed by the Commission or an Agreement State to perform such services.
14. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
  15. When performing tests at temporary job sites, the authorized user shall not leave the moisture/density gauge unattended. Upon completion of tests the device shall be locked in the licensee's vehicle or a secure building to prevent unauthorized use, loss, or theft.
  16. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed under the license.
  17. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
  18. Each portable nuclear gauge shall have a lock or outer locked container designed to prevent unauthorized or accidental removal of the sealed source from its shielded position. The gauge or its container must be locked when in transport, storage, or when not under the direct surveillance of an authorized user.
  19. Any cleaning, maintenance, or repair of the gauge(s) that requires removal of the source rod shall be performed only by the manufacturer or by other persons specifically licensed by the Commission or an Agreement State to perform such services.
  20. The licensee may not possess and use materials authorized in Items 6, 7, and 8 until:
    1. The licensee has constructed the facilities and obtained the equipment described in the application and supporting documentation; and
    2. The U. S. Nuclear Regulatory Commission, Region III, ATTN: Chief, Materials Licensing Section, 801 Warrenville Road, Lisle, IL 60532-4351 has been notified that activities authorized by the license will be initiated.
  21. Within 30 days of the date of a decision not to complete the facility, acquire equipment, or possess and use authorized material, the licensee must notify the Commission in writing, of the decision.

COPY

MATERIALS LICENSE  
SUPPLEMENTARY SHEET

License Number

48-26614-01


Docket or Reference Number

030-33693

Amendment No. 01

22. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- A. Application dated October 7, 1994; and
- B. Letters dated December 5, 1994, December 9, 1994, April 28, 1997 and June 2, 1997.



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date JUN 05 1997

By

*Michael F. Webb*

Nuclear Materials Licensing Branch, Region III

COPY



BETWEEN:

License Fee Management Branch, ARM  
and  
Regional Licensing Sections

(FOR LFMS USE)  
INFORMATION FROM LTS

Program Code: 03121  
Status Code: 0  
Fee Category: 3P  
Exp. Date: 20041231  
Fee Comments:  
Decom Fin Assur Req'd: N

52

LICENSE FEE TRANSMITTAL

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: CGC INCORPORATED  
Received Date: 970505  
Docket No: 3033693  
Control No.: 302598  
License No.: 48-26614-01  
Action Type: Amendment

2. FEE ATTACHED

Amount: 300  
Check No.: 3574

3. COMMENTS

Signed  
Date

D. Hershey  
5-7-97

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / K)

1. Fee Category and Amount:

3P \$300

2. Correct Fee Paid. Application may be processed for:

Amendment ☒  
Renewal ☐  
License ☐

3. OTHER

Signed  
Date

SC 5/12/97

MAY 19 1997

Log	May 4 711
Remitter	
Check No.	5574
Amount	\$300
Fee Category	3P
Type of Fee	AmD
Date Check Rec'd	5/12/97
Date Completed	5/12/97
By:	SC

1997 MAY 12 PM 4:45

# CGC, Inc.

Construction • Geotechnical  
Consulting Engineering/Testing

April 28, 1997

US Nuclear Regulatory Commission  
Region III  
801 Warrenville Road  
Lisle, IL 60532

Attention: Material Licensing Branch

Re: Application for Amendment of Materials License  
License No. 48-26614-01

To Whom It May Concern:

CGC, Inc. is applying for several amendments to our Materials License. They pertain to the following areas:

## **Part 2 - Licensee Address**

The new address for the licensee is as follows:

3011 Perry Street  
Madison, WI 53713

## **Part 8 - Maximum Amount that Licensee May Possess**

CGC requests an increase of the maximum amount that we may possess under this license from 12 to 15 of each sealed source.

## **Part 10 - Licensee's Storage Facility**

We request that our storage facility address be changed from 1409 Emil Street, Madison, WI to 3011 Perry Street, Madison, WI. Also, we would like to add our Milwaukee office as a permanent storage facility. The address is as follows:

205 W. Galena, Suite 100  
Milwaukee, WI 53212

ltr 74-nrc.kjl

3011 Perry Street, Madison, WI 53713  
Telephone: 608/288-4100  
FAX: 608/288-7887

*Pm: 5-2-97*

RECEIVED  
MAY 05 1997  
REGION III

MAY 05 1997

302598

US Nuclear Regulatory Commission  
Region III  
Page 2  
April 28, 1997


\*\*\*\*\*

Enclosed you will find a copy of our license (with the proposed changes) and sketches of the new storage facilities at 3011 Perry Street and 205 W. Galena. Also enclosed is a check for the license application for amendment fee.


If you have any questions, please contact us.

Sincerely,

CGC, INC.




Kim J. Lewis  
Materials Laboratory Supervisor/RSO




Michael N. Schultz, P.E.  
President  
KJL/mes/MNS

FLOOR PLAN @  
3011 PERRY ST.

- ① TROXLER STORAGE ROOM
- ② MISL. STORAGE ROOM
- ③ OFFICES WITH WORK STATION/DESK
- ④ COPIER OR FILE ROOM

 TROXLER STORAGE AREA


 DESK LOCATION

- OFFICE NOTES
- 1 2x4 STUD WALL
  - 2 8' 1/2" DW T.
  - 3 COAT LATER
  - 4 VINYL BASE
  - 5 ACOUSTIC CEILING
  - 6 CARPET FLOOR
  - 7 INSUL FRONT

EAST  
TERR




NO SCALE





FLOOR PLAN OF THE PREMISES  
(205 W. GALENA - SUITE 100)

NUCLEAR GAUGE  
STORAGE ROOM  
(9.5' x 10')



First Floor Plan

SCALE: 1" = 10'



## MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 39, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

## Licensee

1. CGC, Inc.

3. License number 48-26614-01

2. ~~1409 Emil Street~~  
Madison, WI 53713

3011 PERRY ST

4. Expiration date December 31, 1999

5. Docket or  
Reference No. 1030-336936. Byproduct, source, and/or  
special nuclear material7. Chemical and/or physical  
form8. Maximum amount that licensee  
may possess at any one time  
under this license

A. Cesium-137

A. Sealed source  
(Troxler Dwg.  
No. A-102112)A. <sup>15</sup>/<sub>12</sub> sources not  
to exceed 10  
millicuries each

B. Americium-241

B. Sealed source  
(Troxler Dwg.  
No. A-102451)B. <sup>15</sup>/<sub>12</sub> sources not  
to exceed 50  
millicuries each

9. Authorized Use:

A. and B. To be used in Troxler Model 3400 Series moisture/density gauges.

★ ★ ★ ★ ★  
CONDITIONS

10. <sup>3011 PERRY</sup> Licensed material may be used and stored at the licensee's facilities located at <sup>205 W. GALENA, SUITE 100</sup>  
<sup>MILWAUKEE, WI 53212</sup> ~~1409 Emil Street~~, Madison, Wisconsin and at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.

11. The Radiation Safety Officer for this license is Kim J. Lewis.

12. Licensed material shall only be used by, or under the supervision and in the physical presence of, either Kim J. Lewis or Donald W. Arenander or individuals who have successfully completed the manufacturer's training program for gauge users, have been instructed in the licensee's routine and emergency operating procedures and who have been designated by the Radiation Safety Officer.

**MATERIALS LICENSE**  
SUPPLEMENTARY SHEET

License number

48-26614-01

Docket or Reference number

030-33693

13. A. Sealed sources and detector cells shall be tested for leakage and/or contamination at intervals not to exceed 6 months or at such other intervals as specified by the certificate of registration referred to in 10 CFR 32.210.
- B. Notwithstanding Paragraph A of this Condition, sealed sources designed to emit alpha particles shall be tested for leakage and/or contamination at intervals not to exceed 3 months.
- C. In the absence of a certificate from a transferor indicating that a leak test has been made within 6 months prior to the transfer, a sealed source or detector cell received from another person shall not be put into use until tested.
- D. Sealed sources need not be leak tested if:
- (i) they contain only hydrogen-3; or
  - (ii) they contain only a radioactive gas; or
  - (iii) the half-life of the isotope is 30 days or less; or
  - (iv) they contain not more than 100 microcuries of beta and/or gamma emitting material or not more than 10 microcuries of alpha emitting material; or
  - (v) they are not designed to emit alpha particles, are in storage, and are not being used. However, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source or detector cell shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
- E. The leak test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(b)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. The report shall be filed within 5 days of the date the leak test result is known with the U.S. Nuclear Regulatory Commission, Region III, ATTN: Chief, Nuclear Materials Safety Branch, 801 Warrenville Road, Lisle, Illinois 60532-4351. The report shall specify the source involved, the test results, and corrective action taken.

**MATERIALS LICENSE  
SUPPLEMENTARY SHEET**

License number

48-26614-01

Docket or Reference number

030-33693

- F. The licensee is authorized to collect leak test samples for analysis by Troxler Electronic Laboratories. Alternatively, tests for leakage and/or contamination may be performed by persons specifically licensed by the Commission or an Agreement State to perform such services.
14. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
15. When performing tests at temporary job sites, the authorized user shall not leave the moisture/density gauge unattended. Upon completion of tests the device shall be locked in the licensee's vehicle or a secure building to prevent unauthorized use, loss, or theft.
16. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed under the license.
17. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
18. Each portable nuclear gauge shall have a lock or outer locked container designed to prevent unauthorized or accidental removal of the sealed source from its shielded position. The gauge or its container must be locked when in transport, storage, or when not under the direct surveillance of an authorized user.
19. Any cleaning, maintenance, or repair of the gauge(s) that requires removal of the source rod shall be performed only by the manufacturer or by other persons specifically licensed by the Commission or an Agreement State to perform such services.
20. The licensee may not possess and use materials authorized in Items 6, 7, and 8 until:
1. The licensee has constructed the facilities and obtained the equipment described in the application and supporting documentation; and
  2. The U. S. Nuclear Regulatory Commission, Region III, ATTN: Chief, Materials Licensing Section, 801 Warrenville Road, Lisle, IL 60532-4351 has been notified that activities authorized by the license will be initiated.
21. Within 30 days of the date of a decision not to complete the facility, acquire equipment, or possess and use authorized material, the licensee must notify the Commission in writing, of the decision.

MATERIALS LICENSE  
SUPPLEMENTARY SHEET


License number

48-26614-01

Docket or Reference number

030-33693

22. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated October 7, 1994; and
- B. Letters dated December 5, 1994 and December 9, 1994.



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date 12/19/94

By

James Mullawer  
Materials Licensing Section, Region III



JUN 05 1997

Michael N. Schultz, P.E.  
Principal  
CGC, Inc.  
3011 Perry Street  
Madison, WI 53713

Dear Mr. Schultz:

Enclosed is Amendment No. 01 to your NRC Material License No. 48-26614-01 in accordance with your request.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region III office at (630) 829-9887 so that we can provide appropriate corrections and answers.

Please be advised that your license expires at the end of the day, in the month, and year stated in the license. Unless your license has been terminated, you must conduct your program involving byproduct materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Notify NRC, in writing, within 30 days:
  - a. When the Radiation Safety Officer permanently discontinues performance of duties under the license or has a name change; or
  - b. When the mailing address listed on the license changes. (No fee is required if the location of byproduct material remains the same.)

302598

3. In accordance with 10 CFR 30.36(b) and/or license condition, notify NRC, promptly, in writing, and request termination of the license when a decision is made to terminate all activities involving materials authorized under the license.
4. Request and obtain a license amendment before you:
  - a. Change Radiation Safety Officers;
  - b. Order byproduct material in excess of the amount, or radionuclide, or form different than authorized on the license;
  - c. Add or change the areas of use or address or addresses of use identified in the license application or on the license; or
  - d. Change ownership of your organization.
5. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date of your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of byproduct material after your license expires is a violation of NRC regulations. A license will not normally be renewed, except on a case-by-case basis, in instances where licensed material has never been possessed or used.

In addition, please note that NRC Form 313 requires the applicant, by his/her signature, to verify that the applicant understands that all statements contained in the application are true and correct to the best of the applicant's knowledge. The signatory for the application should be the licensee or certifying official rather than a consultant.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions, and representations made in your license application and supplemental correspondence with NRC will result in enforcement action against you. This could include issuance of a notice of violation, or imposition of a civil penalty, or an order suspending, modifying or revoking your license as specified in the General Statement of Policy and Procedure for NRC Enforcement Actions. Since serious consequences to employees and the public can result from failure to comply with NRC

M. Schultz

-3-

requirements, prompt and vigorous enforcement action will be taken when dealing with licensees who do not achieve the necessary meticulous attention to detail and the high standard of compliance which NRC expects of its licensees.

Sincerely,

Original Signed By  
Michael F. Weber  
Nuclear Materials Licensing Branch

License No.: 48-26614-01  
Docket No.: 030-33693

Enclosure: Amendment No. 01

DOCUMENT NAME: M:\03033693.CL7

To receive a copy of this document, indicate in the box: "C" = Copy without attachment/enclosure "E" = Copy with attachment/enclosure "N" = No copy

OFFICE	DNMS/RIII	C							
NAME	MFWeber:brt								
DATE	06/ 4 /97								

OFFICIAL RECORD COPY

# CGC, Inc.

Construction • Geotechnical  
Consulting Engineering/Testing

June 2, 1997

US Nuclear Regulatory Commission  
Region III  
801 Warrenville Road  
Lisle, IL 60532

Attention: Mr. Mike Weber  
Material Licensing Branch

Re: Mail Control No. 302598  
Application for Amendment of Material License  
License No. 48-26614-01

Dear Mr. Weber:

This letter is in response to a telephone conversation you had with Mr. John Thorton, of Troxler Electronics, on May 28, 1997. It is our understanding you requested clarification on three items from page 17 of the Regulation Licensing Guide (see attached sheet) in order to process our amendment request. Our response regarding those items are listed below:

1. Both proposed facilities currently exist;
2. Both proposed facilities are in an industrial park setting; and
3. As indicated on the sketches that were included in our amendment request, both rooms are separate and dedicated to storage of the gauges and associated equipment. The Perry Street facility is approximately 6 ft x 11 ft and the West Galena Street facility is 9.5 ft x 10 ft. Each is locked with access only to those authorized.


Thank you for your attention to this matter.

Sincerely,

CGC, INC.



Kim J. Lewis, R.S.O.  
Materials Laboratory Supervisor



Michael N. Schultz, P.E.  
Principal/Consulting Professional

May\usnucler.kjl

3011 Perry Street, Madison, WI 53713  
Telephone: 608/288-4100  
FAX: 608/288-7887

**RECEIVED**  
**JUN 03 1997**  
**REGION III**

**JUN 08 1997**

tended under the constant surveillance and immediate control of the licensee. NRC Information Notice (IN) 93-18, "Portable Moisture-Density Gauge User Responsibilities During Field Operations," provides additional information on these matters and is reproduced in Appendix E to this guide.

Note also that 10 CFR 20.1003 defines a restricted area as "an area, access to which is limited by the licensee for the purpose of protecting individuals against undue risks from exposure to radiation and radioactive materials. Restricted area does *not include areas used as residential quarters* (emphasis added), but separate rooms in a residential building may be set apart as a restricted area." Thus, it is not acceptable to store portable gauges in motel rooms or similar locations.

In response to Item 9, provide the following information:

→ 1. For each proposed permanent facility listed in Item 3 of NRC Form 313, specify whether the proposed facility currently exists, is under construction, or is planned for future construction. If the facility is under construction or planned for future construction, include the estimated completion date.

→ 2. Describe the general location of each proposed permanent facility listed in Item 3 (e.g., located in an industrial park, an office complex, a private residence) and its current use. If any proposed permanent facility is a private residence, confirm that the use of licensed material does not conflict with local codes or zoning laws, provide diagrams of the facility that include the building, the proposed restricted area or areas and adjacent areas, including above and below the restricted areas; provide commitments that restricted areas do not include residential quarters; and explain how radiation levels in unrestricted areas will be controlled and monitored to comply with 10 CFR 20.1301.

→ 3. Describe the storage locations at each address listed in Item 3 of the application and submit a diagram showing where the gauge will be stored when not at field locations. For example, is the storage location a closet? a separate room dedicated to storage of gauges only? a garage?



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION III  
801 WARRENVILLE ROAD  
LISLE, ILLINOIS 60532-4351  
630-829-9887 (phone), 630-515-1259 (fax)

CONVERSATION RECORD

TIME

3:00 pm

DATE

5/28/97

NAME OF PERSON(S) CONTACTED

ORGANIZATION

TELEPHONE NO.

John Thorton

Troxler

847-695-0900

SUBJECT

Amendment request - CGC - Control No. 302598.

EXPLANATION

Mr. Thorton asked about the status of the amendment. After a quick scan of the amendment request, I indicated that the licensee needed to address Item 9 of DG-0008 in more detail.

Mr. Thorton indicated that he would forward this information to the licensee.

ACTION REQUIRED


Wait for response.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Michael F. Weber

|  |

5/28/97



UNITED STATES  
NUCLEAR REGULATORY COMMISSION

REGION III  
801 WARRENVILLE ROAD  
LISLE, ILLINOIS 60532-4351

May 7, 1997

Kim J. Lewis  
Radiation Safety Officer  
CGC Incorporated  
3011 Perry Street  
Madison, WI 53713

SUBJECT: ACKNOWLEDGEMENT OF CORRESPONDENCE  
(Letter Dated 04/28/97)

Dear Licensee:

In response to your request, we have completed the initial processing, which is an administrative review of your application for a(n):

☐ New License                      ☒ Amendment                      ☐ Renewal  
☐ Termination                      ☐ Auth User (Amendment not required)  
☐ Other \_\_\_\_\_

No administrative deficiencies were identified during this initial review. However, it should be noted that a technical review may identify omissions in the submitted information.

It appears that your request is routine (see 1-3 below, as applicable).

1. New and amendment actions are normally processed within 90 days, unless we find major deficiencies, or policy issues requiring central program office assistance.
2. Renewal actions are normally processed within 180 days, however, under timely filing (before expiration), you may continue to operate under your existing license.
3. Termination actions are normally processed within 90 days, unless confirmatory surveys following decontamination/decommissioning activities are involved.

A copy of your correspondence has been forwarded to our Licensing Fee and Debt Collection Branch (301/415-6097) for approval of the fee category and amount, if required.

If you have a compelling safety or business-related reason for requesting expedited review, please contact the Materials Licensing Branch at (630) 829-9887. We will try to complete your request as soon as practicable. Any correspondence about this request should reference the control number.

Nuclear Materials Support Branch

Mail Control No. 302598  
License No. 48-26614-01