

June 9, 1997

Old Dominion University
ATTN: Paul T. Sutcliffe
Radiation Safety Officer
1300 West 49th Street
Norfolk, Virginia 23529-0566

Dear Mr. Sutcliffe:

Thank you for your letter of May 9, 1997, informing the NRC of leak test results which exceeded the regulatory limit for removable contamination of .005 microcuries. We have reviewed your letter and will examine the incident and implementation of your corrective actions during a future inspection.

We appreciate your cooperation in this matter.

Sincerely,

(original signed by
T. R. Decker)

Thomas R. Decker, Acting Chief
Nuclear Materials Licensing and Inspection
Branch 1
Division of Nuclear Materials Safety

Docket No. 030-16045
License No. 45-09599-03

cc: Commonwealth of Virginia

Distribution:

PUBLIC

RII Docket File, DNMS

OFFICE	RII-DNMS					
SIGNATURE	<i>TR Decker</i>					
NAME	AR Jones					
DATE	06 / 9 / 97	06 / / 97	06 / / 97	06 / / 97	06 / / 97	06 / / 97
COPY?	YES (NO)	YES (NO)	YES NO	YES NO	YES NO	YES NO

OFFICIAL RECORD COPY

DOCUMENT NAME: G:\DNMS\MLIB\ODU.ARJ

9706180228 970609
PDR ADOCK 03016045
C PDR



ML21