

MAR 11 1986

Brighton Analytical, Inc.
ATTN: Gary J. Wagner
President
1576 Alloy Parkway
Highland, MI 48031

Gentlemen:

Enclosed is your NRC License Number 12-24661-01 in accordance with your request.

Please note your leak test condition. The device registry sheet on file with the U.S. NRC recommends a three year leak test interval. Your application states you will perform a leak test at 6 month intervals. Conditions 16. and 17. authorizes a three year leak test interval. This will allow you to perform leak tests at 3 year intervals if you wish, rather than the 6 month intervals you've committed to in your application.

Please review the enclosed document carefully and be sure that you understand all conditions. You must conduct your program involving radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, note that you must:

1. Operate in accordance with NRC regulations 10 CFR Part 19, "Notices, Instructions and Reports to Workers; Inspections," 10 CFR Part 20, "Standards for Protection Against Radiation," and other applicable regulations.
2. Possess radioactive material only in the quantity and form indicated in your license.
3. Use radioactive material only for the purpose(s) indicated in your license.
4. Notify NRC in writing of any change in mailing address.
5. Request and obtain appropriate amendment if you plan to change ownership of your organization, change locations of radioactive material, or make any other changes in your facility or program which are contrary to your license conditions or representations made in your license application and any supplemental correspondence with NRC. Any amendment request should be accompanied by the appropriate fee specified in 10 CFR Part 170.
6. Submit a complete renewal application with proper fee or termination request at least 30 days before the expiration date on your license. You will receive a reminder notice approximately 90 days before the expiration date. Possession of radioactive material after your license expires is a violation of NRC regulations.

B604040133 B60311
REG3 LIC30
21-24661-01 PDR

7. Request termination of your license if you plan to permanently discontinue activities involving radioactive material prior to your expiration date.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions and representations in your license application will result in enforcement action against you in accordance with the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

If you have any questions or require clarification of any of the above stated information, contact us at (312) 790-5625.

Sincerely,

Original Signed By
Patricia M. Vacherlon
Materials Licensing Section

Enclosure(s):

1. License No. 21-24661-01
2. 10 CFR Parts 19, 20, 30 and 170
3. 10 CFR Part 2, Appendix C

RIII

Pmv
Vacherlon/cm

3/3/86