

**MATERIALS LICENSE**  
SUPPLEMENTARY SHEET

License number  
53-17797-01

Docket or Reference number  
030-13337

Amendment No. 11

Kuakini Medical Center  
347 North Kuakini Street  
Honolulu, Hawaii 96817

In accordance with letter dated July 30, 1985, License No. 53-17797-01 is amended as follows:

Subitem K. and License Condition 20. are amended to read:

- |   |                                     |  |
|---|-------------------------------------|--|
| 6. BYPRODUCT, SOURCE, AND/OR SPECIAL NUCLEAR MATERIAL   | 7. CHEMICAL AND/OR PHYSICAL FORM    | 8. MAXIMUM AMOUNT THAT LICENSEE MAY POSSESS AT ANY ONE TIME UNDER THIS LICENSE |
| K. Iodine 125   | K. Sealed source (AECL Model C-235) | K. Not to exceed 800 millicuries per source                                    |
| 9. Authorized use   |                                     |  |
| K. For use in an AECL Model C-236 source holder in a Model JMV-1 or in a Model SPSHA 110 or in a Model SPSHA 220 Bone Mineral Analyzer for bone mineral analysis. |                                     |  |

CONDITIONS

20. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated August 26, 1983, and letters dated December 7, 1983, January 19, 1984, October 22, 1984, and July 30, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

AUG 20 1985

Date \_\_\_\_\_

By Beth A. Rieckinger  
Health Physicist (Licensing)  
Nuclear Materials Safety Section  
Region V

8510230092 850820  
REG5 LIC30  
53-17797-01 PDR

ML 50