

OCT 02 1985

License No. 29-02575-01  
Docket No. 030-02449  
Control No. 104114

Muhlenberg Hospital  
ATTN: David M. Ridgway, President  
Randolph Road  
Plainfield, New Jersey 07061

Gentlemen:

Please find enclosed an amendment to your NRC Material License.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the Region I Material Licensing Section, (215) 337-5239, so that we can provide appropriate corrections and answers.

Please be advised that you must conduct your program involving licensed radioactive materials in accordance with the conditions of your NRC license, representations made in your license application, and NRC regulations. In particular, please note the items in the enclosed, "Requirements for Materials Licensees."

Please note we have separated out each Model No. authorized for the Norland and Lunar bone mineral analyzers in order to assure the total activity registered for the source model by the manufacturer for use in the bone mineral analyzer is not exceeded. Additionally, please note that the Lunar Model DP3 bone mineral analyzer is the only device authorized for the Amersham, NEN and Gulf Nuclear Model Gadolinium-153 sources at this time. Lunar Radiation Corporation has designated these 3 sources as the GD-Series. Therefore our use of GD-Series in item 7N authorizes the Amersham GDC 10413 and NEN NER-430 without listing them specifically.

Since serious consequences to employees and the public can result from failure to comply with NRC requirements, the NRC expects licensees to pay meticulous attention to detail and to achieve the high standard of compliance which the NRC expects of its licensees.

You will be periodically inspected by NRC. A fee may be charged for inspections in accordance with 10 CFR Part 170. Failure to conduct your program safely and in accordance with NRC regulations, license conditions, and representations made in your license application and supplemental correspondence with NRC will result in prompt and vigorous enforcement action against you. This could include issuance of a notice of violation, or in case of serious violations, an imposition of a civil penalty or an order suspending, modifying or revoking your license as specified in the General Policy and Procedures for NRC Enforcement Actions, 10 CFR Part 2, Appendix C.

OFFICIAL RECORD COPY

ML 29-02575-01/LTR - 0001.0.0  
09/30/85

8510230321 851002  
REG1 LIC30  
29-02575-01 PDR

ML10

We wish you success in operating a safe and effective licensed program.

Sincerely,

Original Signed By:  
Jenny M. Johansen


Jenny M. Johansen, M.S.  
Nuclear Materials Safety Section B  
Division of Radiation Safety  
and Safeguards

Enclosures:

1. Amendment No. 50
2. Requirements for Materials Licensees

RI:DRSS  
Johansen/mmb

10/1/85



OFFICIAL RECORD COPY

ML 29-02575-01/LTR - 0002.0.0  
09/30/85