

## MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee		
1. Department of the Army Academy of Health Sciences (AHS) Preventive Medicine Division HSHA-IPM	3. License number	42-01368-04
2. Radiation Protection Officer Fort Sam Houston, Texas 78234	4. Expiration date	August 31, 1990
	5. Docket or Reference No.	030-28823
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Any byproduct material with Atomic Numbers 3 through 83, inclusive	A. Sealed sources	A. Not to exceed 1 millicurie per sealed source or foil
B. Americium-241	B. Sealed sources	B. Not to exceed 1 millicurie per sealed source
9. Authorized use:		
A. and B. For use in academic instruction.		

## CONDITIONS

10. Licensed material shall be used only at the Department of the Army's Academy of Health Sciences facilities, Fort Sam Houston, Texas.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and Part 20, "Standards for Protection Against Radiation."

8510230341 850826  
REG4 LIC30  
42-01368-04 PDR

**MATERIALS LICENSE**  
SUPPLEMENTARY SHEET

License number  
42-01368-04

Docket or Reference number  
030-28823

12. Licensed material shall be used by, or under the supervision of, individuals designated by Fort Sam Houston Academy of Health Sciences Radiation Control Committee.
13. A. (1) Each sealed source containing licensed material, other than hydrogen-3, with a half-life greater than 30 days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed 6 months, except that each source designed for the purpose of emitting alpha particles shall be tested at intervals not to exceed 3 months. In the absence of a certificate from a transferor, indicating that a test has been made within 6 months prior to the transfer, a sealed source received from another person shall not be put into use until tested.
- (2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.
- (3) Except for alpha sources, the periodic leak test required by this condition does not apply to sealed sources that are stored and not being used. The sources excepted from this test shall be tested for leakage prior to any use or transfer to another person unless they have been leak tested within 6 months prior to the date of use or transfer.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the test with the U. S. Nuclear Regulatory Commission, Region IV, 611 Ryan Plaza Dr, Suite 1000, Arlington, Texas 76011, describing the equipment, the test results, and the corrective action taken.



**MATERIALS LICENSE**  
SUPPLEMENTARY SHEET

License number  
42-01368-04

Docket or Reference number  
030-28823

- D. The licensee is authorized to collect and analyze leak test samples in accordance with the procedures described in the licensee's application dated November 8, 1984. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Commission or an Agreement State to perform such services.
14. Licensed material shall not be used in or on human beings.
15. Sealed sources containing licensed material shall not be opened.
16. The licensee is authorized to hold radioactive material with a physical half-life of less than 65 days for decay-in-storage before disposal in ordinary trash provided:
- A. Radioactive waste to be disposed of in this manner shall be held for decay a minimum of 10 half-lives.
- B. Prior to disposal as normal waste, radioactive waste shall be monitored to determine that its radioactivity cannot be distinguished from background with typical low-level laboratory survey instruments. All radiation labels will be removed or obliterated.
17. The licensee may transport licensed material or deliver licensed material to a carrier for transport in accordance with the provisions of Title 10, Code of Federal Regulations, Part 71, "Packaging and Transportation of Radioactive Material."
18. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated November 8, 1984. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date AUG 26 1985

Original Signed By  
By Jack E. Whitten  
Nuclear Materials Safety Section  
Region IV  
Arlington, Texas 76011

Official Record Copy

11 MC40