

File

NOV 6 1985

McAuley Clinic Corporation  
ATTN: Jeffrey Susi  
President  
5333 McAuley Drive  
Ann Arbor, MI 48106

Gentlemen:

We have reviewed your letter dated September 18, 1985 requesting an amendment and find that we will need additional information as follows:

1. Please submit a copy of the user instructions supplied by Lunar.
2. Describe the methods you will use to prevent unauthorized use/removal of the Gadolinium-153 source.
3. Describe the type of personnel monitoring you will use during a source exchange to assure the person performing the source exchange is not over exposed.
4. Describe what disposition will be made of depleted sources.

We will continue our review of your application upon receipt of this information. Please reply in duplicate, within 30 days, and refer to Control Number 79852.

Sincerely,

Original Signed By  
Patricia M. Vacherlon  
Materials Licensing Section

8604040442 860122  
REG3 LIC30  
21-24447-01 PDR

RLII  
P.M.V.  
Vacherlon/cm

10/28/85

11/05/85