

MATERIALS LICENSE

Amendment No. 02

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee		In accordance with application dated October 17, 1985	
1. Norprop Films Division of Norchem, Inc.		3. License number 12-20060-01 is amended in its entirety to read as follows:	
2. 601 E. Lake Street Streamwood, IL 60103		4. Expiration date	November 30, 1990
		5. Docket or Reference No.	030-17915
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license	
A. Strontium-90	A. Sealed sources (Amersham Corp. Model No. VZ-255a)	A. 3 sources not to exceed 15 millicuries each	
B. Promethium-147	B. Sealed sources (Amersham Corp. Model No. VZ-0095)	B. 3 sources not to exceed 500 millicuries each	
9. Authorized Use			
A. To be used in an Amersham Corp. Model VZ-0257 source holder, TIAM 11-91 source container in Hartmann and Braun custom gauge for thickness measurements.			
B. To be used in an Amersham Corp. Model ES88 source holder, TIAM 11-91 source container in Hartmann and Braun custom gauge for thickness measurements.			

CONDITIONS

10. Licensed material shall be used only at the licensee's facilities located at
601 E. Lake Street, Streamwood, Illinois.
11. The licensee shall comply with the provisions of Title 10, Chapter 1, Code of Federal
Regulations, Part 19, "Notices, Instructions and Reports to Workers; Inspections" and
Part 20, "Standards for Protection Against Radiation."
12. Licensed material shall be used by, or under the supervision of, Kenneth Harris or
John R. Kotal.

8602240485 851122
REG LIC30
12-20060-01 PDRml30 BL COPY 2
NOV 22 1985

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number 12-20060-01

Docket or Reference number 030-17915

Amendment No. 02

13. A. (1) Each sealed source containing licensed material, other than Hydrogen 3, with a half-life greater than thirty days and in any form other than gas shall be tested for leakage and/or contamination at intervals not to exceed six months. In the absence of a certificate from a transferor indicating that a test has been made within six months prior to the transfer, a sealed source received from another person shall not be put into use until tested.
- (2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.
- B. The test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. The test sample shall be taken from the sealed source or from the surfaces of the device in which the sealed source is permanently mounted or stored on which one might expect contamination to accumulate. Records of leak test results shall be kept in units of microcuries and maintained for inspection by the Commission.
- C. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the licensee shall immediately withdraw the sealed source from use and shall cause it to be decontaminated and repaired or to be disposed of in accordance with Commission regulations. A report shall be filed within five (5) days of the test with the U. S. Nuclear Regulatory Commission, Region III, 799 Roosevelt Road, Glen Ellyn, Illinois 60137, describing the equipment involved, the test results, and the corrective action taken.
- D. The licensee is authorized to collect leak test samples in accordance with the procedures described in the licensee's application dated October 17, 1985 for analysis by Health Physics Associates, Ltd. Alternatively, leak test samples may be collected and/or analyzed by other persons specifically authorized by the Commission or an Agreement State to perform such services.
14. Sealed sources containing licensed material shall not be opened or removed from their respective source holders by the licensee.
15. Installation, relocation, removal from service, maintenance, repair, and initial radiation survey of devices containing licensed material and installation, replacement, and disposal of sealed sources containing licensed material used in devices shall be performed only by the manufacturer or by other persons specifically authorized by the Commission or an Agreement State to perform such services.
16. The licensee may transport licensed material or deliver licensed material to a carrier for transport in accordance with the provisions of Title 10, Code of Federal Regulations, Part 71, "Packaging of Radioactive Material for Transport and Transportation of Radioactive Material Under Certain Conditions."

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number 12-20060-01

Docket or Reference number 030-17915

Amendment No. 02

17. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in applications dated October 29, 1980 and October 17, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.



For the U.S. Nuclear Regulatory Commission

Date NOV 22 1985

Original Signed
By Evelyn R. Matson
Materials Licensing Section, Region III

COPY