



UNIVERSAL CONSULTANTS, INC.

P.O. Box 223  
North Ridgeville, Ohio 44039-0223  
(216) 327-9432

L+L = 20964  
030-22160

U.S.N.R.C.  
Materials Licensing Branch  
Region I  
631 Park Ave.  
King of Prussia, Pa 19406

Re: New Application

Gentlemen:

Enclosed please find a complete application for WARD MAXSON, M.D.  
The following supplementary information is submitted.

1. The emergency procedures will be posted and implemented as necessary. The individuals to be notified and the phone numbers will also be posted.
2. Wipe test analysis, if needed, will be performed in accordance with procedures outlined for Universal Consultants, Inc., NRC license # 34-20327-01.
3. ALARA - As soon as new staff members report for duty and at least annually thereafter, training will be provided for the safe handling of the sources, and the Bone Densitometer. Annually the personnel monitoring results will be reviewed. Consideration will be given to ways of reducing exposure and implementing new procedures.
4. The formal course on Basic Radioisotope Handling Techniques was conducted in accordance with NRC's Policy and Guidance Directive FC 85-1. The specific device instruction will be given during installation. Instruction was given on source exchanging and wipe test procedures during the course.
5. The manufacturer's operating procedure manual will be followed for the safe operation of the equipment and sources.

If you have any questions regarding this application, please do not hesitate to contact Mr. Vincent A. Gargaro with Universal Consultants, Inc., North Ridgeville, Ohio at (216) 327-9432.

Application prepared by: Vincent A. Gargaro  
Vincent A. Gargaro  
Universal Consultants, Inc.

Date

Jan 6, 1986

Application reviewed and approved by:

Ward Maxson

8604070209 860210  
REG2 LIC30  
47-20964-01

PDR

"OFFICIAL RECORD COPY"

204893  
JAN 13 1986