



1282 Alturas Drive – 1282 Alturas Drive – Moscow, ID 83843 – 208-883-2839 – [moscow@anateklabs.com](mailto:moscow@anateklabs.com)  
504 E. Sprague Ave, Suite D – Spokane, WA 99202 – 509-838-3999 – [Spokane@anateklabs.com](mailto:Spokane@anateklabs.com)

---

19 May 2020

**Mail Control Number: 618090**  
**Docket Number : 3038790**  
**License Number : 11-35192-01**  
**Licensee Name : Anatek Labs, Inc**

Michelle Hammond  
Materials Licensing and Decommissioning Branch  
Region IV – US NRC

Re: License Number 11-35192-01

Ms. Hammond:

In response to an inquiry from Dr. Robert Evans, we submit the following information.

Anatek Labs, Inc. has never had a spill or contamination event at any time in our radionuclide laboratory.

Our GM survey meter is a Ludlum Model 3 with a 44-9 probe, calibrated in October 2019. Our wipe samples were analyzed on a Protean GFPC Model 9604.

If you have any questions or need additional information, please do not hesitate to contact us.

Thank you.

Gene Solomon  
QA Officer