

From: [Tran, Frank](#)
To: [Gray, Michael S CIV USARMY CHEMICAL SCHL \(USA\)](#)
Cc: [Atabay, Mahmut A CPT USARMY CHEMICAL SCHL \(USA\)](#)
Subject: Request additional information for NRC License No. 24-32221-01
Date: Friday, May 15, 2020 8:51:00 AM

Dear Mr. Gray:

Based on the authorization for Troxler Electronic Laboratories Model 3400 Series nuclear gauge in the license, we also used the licensing guidance NUREG-1556, Volume 1, Revision 2 in the review of your application. Based on our review, we will need the following.

1) Before using the nuclear portable gauges, the user will have to complete at a minimum the training described in Appendix C, "Criteria for Acceptable Training Courses for Portable Gauge Users", to NUREG-1556, Volume 1, Revision 2. You could provide an alternative procedure if you want.

1) A confirmation that the licensee will use the portable gauges in accordance with their respective sealed source and device registry (SSDR) and will follow the manufacturer/vendor instructions and recommendations for routine and non-routine maintenance for the gauges.

For your reference, a copy of the NUREG-1556, Volume 1, Revision 2 can be found on the NRC's website at <https://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v1/>.

Please provide a response in writing with signature by May 19, 2020. You could email the response to frank.tran@nrc.gov (no need to mail to original copy to me). If you have any question, please contact me at 630-829-9623 or reply to this email.

Thank you,

Frank Tran