

OCT 4 1978

The Ohio State University  
ATTN: Dr. Henry Cramblatt, M.D.  
Vice President for  
Medical Affairs  
Columbus, OH 43210

License No. 34-00293-02  
License No. 34-00293-12  
License No. SNM-917  
License No. SNM-732  
License No. SUD-846

Gentlemen:

Thank you for your letter dated August 21, 1978, informing us of the steps you have taken to correct the noncompliance identified in our letter of August 3, 1978. We have no further questions regarding your response to noncompliance item 1.

Our position on noncompliance item 2 continues to be as stated in our letter dated August 3, 1978 namely, all survey meters used to measure dose or contamination from NRC licensed materials are required to be calibrated annually as stated in item 15, page 8 of your letter dated August 8, 1974. If your intention is to require annual calibration of only those instruments used by the Radiation Safety Office you must amend your license to state the change in item 15.

Based on your letter of August 21, 1978 and as a result of a telephone conversation on September 29, 1978 with Mr. J. Dare of your staff, we understand that your required leak test method is capable of detecting .005 microcuries of removable alpha contamination. Further, we understand that records of these tests are recorded in units of microcuries. Consequently, this resolves the matter discussed in Appendix B of our letter. We have no further questions regarding this matter.

We have no further questions on the steps you have taken to correct the noncompliance items 2 and 3 (former Appendix B). We will examine your corrective actions during a future inspection.

Your cooperation with us is appreciated.

Sincerely,

7810110179

A. B. Davis, Chief  
Fuel Facility and Materials  
Safety Branch

cc: J. Dare, OSU Rad. Safety

OFFICE ►	Officer <i>Central Files</i>	RIII	RIII	RIII	RIII
SURNAME ►	Reproduction Unit NRC 20b	Sreniawski/dec	Pagliari	<i>DAVIS</i>	Keppler
DATE ►	PDR	Axelson	10/3/78		
	NSIC				