

P

MAY 7 1987

Tri-County Radiologists
ATTN: R. B. Sneider, D.O.
Radiation Safety Officer
1544 Lakewood
Bloomfield Hills, MI 48013

Gentlemen:

We have reviewed your application dated March 23, 1987 requesting an amendment to NRC License Number 21-17974-01 and find that we will need additional information as follows:

1. In regards to authorizing J. David Cole, M.D. on your license, it appears that he was authorized on a license issued by the State of Illinois. Please be aware that the State of Illinois does not regulate byproduct material, but that this material is regulated by the NRC. The materials listed on your license are byproduct material. Therefore, Dr. Cole must submit the training and experience received in handling byproduct material. To determine the type and amount of training and experience Dr. Cole will need to be authorized on your license, please refer to the enclosed NRC regulation, 10 CFR Part 35, Subpart J.
2. In your application you stated that you will purchase bone mineral analyzers from one or more of several manufacturers. Please indicate which devices and sealed sources you wish to be authorized to use.
3. Please confirm that the reference made in Item No. 23 to factory engineer is equivalent to the device manufacturer.
4. In regard to transporting the devices using iodine-125 sealed sources for use and storage at your other facilities and other licensees specifically authorized to receive them, please submit the following:
 - a. Confirmation that the source will be brought into the facility by the applicant and removed when the applicant departs. (Requested only if the facility is not under the control of the applicant.)
 - b. Description of quality assurance checks performed on the device to verify proper operation prior to use at each of the facilities requested.

8801210360 870727
REG3 LIC30
21-17974-01 PDR

Tri-County Radiologists

2

If you have any questions or require clarification on any of the information stated above, you may contact us at (312) 790-5625.

We will continue our review of your application upon receipt of this information. Please reply in duplicate, within 30 days, and refer to Control Number 83217.

Sincerely,

Original Signed By
Cassandra Frazier
Materials Licensing

Enclosures: As stated

RIII

cf
Frazier/pd
5/7/87