



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
631 PARK AVENUE
KING OF PRUSSIA, PENNSYLVANIA 19406

Docket No. 030-03195

OCT 12 1978

Michael F. Dimun, M. D.
135 E. Mall Plaza
Carnegie, Pennsylvania 15106

Gentlemen:

Subject: Telephone Inquiry in Strontium-90 Source

This refers to a telephone inquiry by Ms. M. Campbell of this office with Mrs. Taylor of your staff on September 29, 1978.

This inquiry was an examination of activities authorized by License No. 37-13604-01, as they relate to the strontium-90 source used in treatment of eye diseases.

From this discussion, it is our understanding:

1. that this source is kept in secure storage in an unoccupied area.
2. that it is in active use as part of your practice.
3. that it was last leak tested about six months ago and another leak test is anticipated.

In accordance with Section 2.790 of the NRC's "Rules of Practice," Part 2, Title 10, Code of Federal Regulations, a copy of this letter will be placed in the Public Document Room. No reply to this letter is required; however, should you have any questions concerning this inquiry, we will be pleased to discuss them with you.

Your cooperation with us is appreciated.

Sincerely,

Robert O. McClintock
Robert O. McClintock, Chief
Materials Radiological Protection
Section

7810200187