

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number

13-16418-01

Docket or Reference number

030-10998

Amendment No. 04

OA Laboratories and Research, Inc.
1437 Sadlier Circle, W. Drive
Indianapolis, IN 46239

In accordance with letter dated May 7, 1987, License Number 13-16418-01 is amended as follows:

Condition 17.A. is amended to read:

17. A. (1) The source(s) specified in Item(s) 7.A shall be tested for leakage and/or contamination at intervals not to exceed 3 years. Any source received from another person which is not accompanied by a certificate indicating that a test was performed within 6 months before the transfer shall not be put into use until tested.

8801220118 870813
REG3 LIC30
13-16418-01 PDR

For the U.S. Nuclear Regulatory Commission

Date August 13, 1987

Original Signed
By Patricia M. Vacherlon
Materials Licensing Section, Region III

COPY 21430
P.V.