O A Laboratories and Research, Inc.

May 7, 1987

U.S. Nuclear Regulatory Commission
Materials Licensing Section
Region III
799 Roosevelt Road
Glen Ellyn, Illinois 60137

RE: Request to amend Materials License number 13-16418-01

Materials Licensing Section:

We are requesting to amend section 17.A. of our license so the testing interval will be once every 3 years (36 months) rather than "not to exceed six months" (Tracor Instrument, Notice stamp dated August 20, 1985 - copy attached). According to our last Wipe Test Report the test was performed on 8-16-85 with the next test date being on or before 8-16-88 (copy attached).

The $60.00 amendment fee is enclosed. Please let me know if additional information is needed to have our request granted.

Sincerely yours,

Thomas M. Crawford

Thomas M. Crawford, President
O A Laboratories and Research, Inc.

TMC/mlw

Enclosures

Log May 15 Tue
Submitter
Check No. 18225
Amount $60
Fee Category 30 mat
Type of Fee
Date Check Rec'd.
Date Completed 5/14/87
By:

RECEIVED
MAY 11 1987
REGION III

MAY 11 1987

8801220126 870813
REG3 LIC30
13-16418-01 PDR

CONTROL NO. 83533

1437 Sadlier Circle W. Dr., Indianapolis, Indiana 46239 · (317) 353-9721

Tracor Instruments

NOTICE

Effective November 7, 1983, all Tracor Ni63 electron capture detectors will require wipe testing only once every three years (36 months). In order to take advantage of this new interval, your specific license must be amended to show testing every 36 months. You may find your state or jurisdiction will require more frequent testing.

This extended interval is covered under Amendment 23 of our Specific License 6-1186 and Amendment 1 of our General License 6-3389GL.

Should you have any additional questions or require assistance with this new benefit, please contact our Radiation Officer, Mr. Carlton Lorfing, at (512) 926-2800, Ext. 6013.

/ve

NOTE: In the event that your Ni63 ECD, or the GC with the Ni63 in it, should require service, we must have a copy of your NRC license in our files .... be sure to check the expiration date .. and get it renewed if expired.

Amended on April 24, 1984 per The State of Texas.

CONTROL NO. 83533

WIPE TEST REPORT

CUSTOMER O A Laboratories, Inc.

DIVISION

ADDRESS 1437 Sadlier Circle, W. Drive Indianapolis, IN 46239

ATTENTION William E. Oatess

TELEPHONE (317) 353-9721 EXT

DETECTOR TYPE NI63 SERIAL NUMBER 2832

INDIVIDUAL PERFORMING WIPE Edward W. Parkerson

DATE WIPE PERFORMED 8/6/85 8-16-85

TEST RESULTS

[✓] LESS THAN .0005 μCI (NORMAL LIMIT)

[ ] BETWEEN .0005 & .005 μCI (OUTSIDE NORMAL LIMITS)

_____ μCI (CALCULATED)

TRACOR INSTRUMENTS GROUP
6500 TRACOR LANE
AUSTIN, TEXAS 78721

BY [signature] INSPECTOR

NEXT WIPE TEST DATE 8/16/88

112600-0001C

(HH)