

MATERIALS LICENSE
SUPPLEMENTARY SHEET

CORRECTED COPY

License number

20-21326-01

Docket or Reference number

030-20520

Amendment No. 01

Dositec, Inc.
120 Alexander Street
P. O. Box 4744
Framingham, Massachusetts 01701

In accordance with letter dated November 14, 1984, License Number 20-21326-01 is amended as follows:

Items 6., 7., 8., and 9. are amended to read:

6. Byproduct, source, and/or special nuclear material

7. Chemical and/or physical form

8. Maximum amount that licensee may possess at any one time under this license

A. Cesium 137

B. Cesium 137

C. Cesium 137

A. Sealed source

B. Sealed source (J. L. Shepherd and Associates Model 6810)

C. Sealed source (Amersham Model No. X.8)

A. 50 curies

B. 440 curies

C. 130 millicuries

9. Authorized use

A. For use in Technical Operations Model 682 Calibrator for calibration of radiation detection instruments.

B. and C. For use in a J. L. Shepherd and Associates Model No. 89 calibrator for calibration of radiation detection instruments.

Condition 15. is amended to read:

15. Except as specifically provided otherwise by this license, the licensee shall possess and use licensed material described in Items 6, 7, and 8 of this license in accordance with statements, representations, and procedures contained in application dated March 30, 1983, and letters dated June 8, 1983, June 23, 1983, November 14, 1984, and April 24, 1985. The Nuclear Regulatory Commission's regulations shall govern the licensee's statements in applications or letters, unless the statements are more restrictive than the regulations.

8801220304 870806
REC1 LIC30 PDR
20-21326-01

For the U.S. Nuclear Regulatory Commission

Original Signed By:

Thomas K. Thompson

Date AUG 06 1987

By

Nuclear Materials Safety and
Safeguards Branch, Region I
King of Prussia, Pennsylvania 19406

"OFFICIAL RECORD COPY" ML10