

MATERIALS LICENSE

Amendment No. 10
CORRECTED COPY

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 40 and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee		In accordance with letter dated June 22, 1987	
1. Sequoyah Fuels Corporation Kerr-McGee Center		3. License number 35-12636-03 is amended in its entirety to read as follows:	
2. Oklahoma City, Oklahoma 73125		4. Expiration date August 31, 1989	
		5. Docket or Reference No. 030-05948	
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license	
A. Cobalt-60	A. Sealed sources (Technical Operations Model 571 source rod)	A. Not to exceed 15 millicuries per source	
B. Cesium-137	B. Sealed sources (Nuclear Chicago Model RR-1.3)	B. Not to exceed 450 millicuries per source	
C. Cesium-137	C. Sealed sources	C. See item 9.C.	
D. Cesium-137	D. Sealed sources (Technical Operations Model 77302)	D. Not to exceed 165 millicuries per source	

8801220375 870827
REG4 LIC30
35-12636-03 PDR

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

35-12636-03

Docket or Reference number

030-05948

Amendment No. 10

CORRECTED COPY

9. Authorized use

- A. To be used in Technical Operations Model T0/571 calibration unit for instrument calibration.
- B. To be used in Nuclear Chicago Model 506 liquid density gauge.
- C. For possession and use in Ohmart and Texas Nuclear devices which have been evaluated and approved for licensing purposes and authorized for distribution under a license issued by the Nuclear Regulatory Commission or an Agreement State.
- D. For use in Technical Operations Model 773 calibrator for instrument calibrations.

CONDITIONS

- 10. Licensed material shall be used only at the Sequoyah Fuels Corporation facility, Moore, Oklahoma.
- 11. Licensed material shall be used by, or under the supervision of, George P. Sakalosky or Kenneth G. Simeroth.
- 12. A. (1) The sources specified in Items 7.A. and 7.D. shall be tested for leakage and/or contamination at intervals not to exceed 6 months and the sources specified in Items 7.B. and 7.C. shall be tested for leakage and/or contamination at intervals not to exceed 3 years. Any source received from another person which is not accompanied by a certificate indicating that a test was performed within 6 months before the transfer shall not be put into use until tested.

(2) Notwithstanding the periodic leak test required by this condition, any licensed sealed source is exempt from such leak tests when the source contains 100 microcuries or less of beta and/or gamma emitting material or 10 microcuries or less of alpha emitting material.

MATERIALS LICENSE
SUPPLEMENTARY SHEET

License number

35-12636-U3

Docket or Reference number

030-05948

Amendment No. 10

CORRECTED COPY

12. (continued)

- B. Any source in storage and not being used need not be tested. When the source is removed from storage for use or transfer to another person, it shall be tested before use or transfer.
- C. The test shall be capable of detecting the presence of 0.05 microcurie of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie or more of removable contamination, the source shall be removed from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. A report shall be filed within 5 days of the date the leak test result is known with the U.S. Nuclear Regulatory Commission, Region IV, 611 Ryan Plaza Drive, Suite 1000, Arlington, Texas 76011, ATTN: Chief, Nuclear Materials and Emergency Preparedness Branch. The report shall specify the source involved, the test results, and corrective action taken. Records of leak test results shall be kept in units of microcuries and shall be maintained for inspection by the Commission. Records may be disposed of following Commission inspection.
- D. Tests for leakage and/or contamination shall be performed by the licensee or by other persons specifically licensed by the Commission or an Agreement State to perform such services.

13. Sealed sources containing licensed material shall not be opened.

14. Installation, initial radiation survey, relocation, or removal from service of devices containing sealed sources shall be performed only by the licensee or by persons specifically licensed by the Commission or an Agreement State to perform such services. Maintenance and repair of devices and installation, replacement, and disposal of sealed sources shall be performed only by persons specifically licensed by the Commission or an Agreement State to perform such services.

15. The licensee shall conduct a physical inventory every 6 months to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 2 years from the date of each inventory.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**

License number

35-12636-03

Docket or Reference number

030-05498

Amendment No. 10

CORRECTED COPY

16. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.

- A. Application dated July 26, 1984
- B. Letter dated May 1, 1985
- C. Letter dated June 22, 1987

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date

AUG 27 1987

Original signed by
JACK E. WHITTEN

By

Nuclear Materials Licensing Section
Region IV
Arlington, Texas 76011

NUREG

[Handwritten signature]
4/1