

"OFFICIAL RECORD COPY"

FEB 19 1988

34-17732

John Charles Faris, M.D.  
2803 Lyndhurst Avenue  
Winston-Salem, NC 27103

Gentlemen:

SUBJECT: INSPECTION REPORT NO. 32-19457-01/87-01 AND  
REQUEST FOR SUPPLEMENTAL RESPONSE

Thank you for your response of January 12, 1988, to our Notice of Violation issued on December 23, 1987, concerning activities conducted at your facility under NRC License No. 32-19457-01.

Your response to Violation A.2 is satisfactory and meets the requirements of 10 CFR 2.201. However, your response to Violation A.1 is not satisfactory. Violation A.1 concerned a failure to perform a linearity test of your dose calibrator during the first quarter of 1987. Your statement that four tests were performed between 11/86 and 10/87 does not necessarily fulfill the quarterly requirements. We interpret quarterly frequencies as within every three months or as within a calendar quarter. Please respond with a commitment to one of these criteria, as part of your corrective action.

Based on your recent statement that generators are only received dry and only in North Carolina, Violation B is amended to delete reference to 10 CFR 20.205(b) and is restated in the Enclosure as a records violation under your License Condition 14. Please respond to our amended Notice of Violation, accordingly.

We will review implementation of your corrective actions during a future inspection. Should you have any questions concerning this letter, please contact us.

Sincerely,

ORIGINAL SIGNED

BY

W. E. CLINE

William E. Cline  
Nuclear Materials Safety and  
Safeguards Branch  
Division of Radiation Safety  
and Safeguards

Enclosure:  
Notice of Violation (Amended)

bcc w/encl:  
Document Control Desk  
State of NC  
✓ G. Sjoblom, IMNS

RII  
MElliott  
2/4/88

RII  
JPotter  
2/10/88

8803170003 880219  
REU2 LIC30  
32-19457-01 DCD

IE01