

REGISTRY OF RADIOACTIVE SEALED SOURCES AND DEVICES
SAFETY EVALUATION OF DEVICE

NO.: NR-342-D-803-G

DATE: SEP 30 1992

PAGE 1 OF 2

DEVICE TYPE: Thickness Gauge Probe

MODEL: TG-602

MANUFACTURER/DISTRIBUTOR:

Harrel, Inc.
16 Fitch Street
East Norwalk, CT 06855

SEALED SOURCE MODEL DESIGNATION:

Amersham Model AMC.65
Du Pont Merck Model NER-478C

ISOTOPE:

Americum-241

MAXIMUM ACTIVITY:

15 millicuries

LEAK TEST FREQUENCY: 6 Months

PRINCIPAL USE: (D) Gamma Gauge

CUSTOM DEVICE: _____ YES _____ X _____ NO



Concurrence: *[Signature]*