



DEPARTMENT OF THE NAVY
OFFICE OF THE CHIEF OF NAVAL OPERATIONS
2000 NAVY PENTAGON
WASHINGTON, DC 20350-2000

5104
Ser N452/20U132633
17 Aug 2020

Ms. Robin Elliott
U.S. Nuclear Regulatory Commission
Region I, DNMS
2100 Renaissance Blvd, Suite 100
King of Prussia, PA 19406-2713

Dear Ms. Elliott:

SUBJECT: RESCISSION OF LICENSE EXEMPTION REQUEST FOR GAMMA
RADIOGRAPHY SURVEY INSTRUMENT CALIBRATION PERIODICITY

The Department of the Navy (DON) sent a letter dated May 22, 2020 requesting an exemption to Title 10 Code of Federal Regulations Part 34.25 such that radiation survey instruments would be calibrated at twelve month intervals instead of six month intervals. Your email of July 30, 2020 provided a comprehensive request for additional information in order to complete the review of the exemption request. The DON has carefully deliberated, and determined that continued pursuit of the exemption is not advantageous at this time. This letter rescinds the original request for exemption.

If you have additional questions, please do not hesitate to contact me via telephone at (703) 695-2020 or through electronic mail at anthony.s.williams4@navy.mil.

Sincerely,

A. S. WILLIAMS
Captain, Medical Services Corps,
United States Navy
Executive Secretary
National Radiation Safety Committee

Copy to: Naval Sea Systems Command (SEA 04N, SEA 04ND)
Naval Sea Systems Command Detachment, Radiological Affairs Support Office