

From: [Richard Hallahan](#)
To: [Wardrobe, Leonardo](#)
Subject: [External_Sender] Missing Leak Test
Date: Thursday, January 28, 2021 3:16:19 PM
Attachments: [image001.png](#)
[image002.png](#)
[image003.png](#)
[image004.png](#)
[image005.png](#)
[image006.png](#)
[20210128145225.pdf](#)

Mr. Wardrobe,

We had to go back to Troxler as they wrote in the wrong Gauge # on the leak test report. They have re-sent the report with the corrected gauge #. Please see attached.

Thank you,

Richard T Hallahan, PE

Senior Vice President

Office (914) 945-9010 ext 323

Cell (917) 613-2265



47 Hudson Street, Bldg A

Ossining, NY 10562

www.ATANEConsulting.com



Before printing this e-mail please consider if it is necessary to do so: The environment concerns us all.

CONFIDENTIALITY NOTICE TO RECIPIENT:

This message (including any attachments) contains confidential information intended for a specific recipient and purpose, and is protected by law. If you are not the intended recipient, you should delete this message and immediately contact the sender. Any retention, dissemination, disclosure, copying, or distribution of this message, or the taking of any action based on it, may be unlawful and is strictly prohibited. If you have received this message in error, please notify the sender and delete it immediately, retaining no copies.

**Troxler Electronic Laboratories, Inc.**

3008 Cornwallis Rd., P.O. Box 12057

Research Triangle Park, NC 27709

Tel: (877) 876-9537 Fax: (866) 391-2759

License: NC 032-0182-1

NASIR KHAN
HAKS MATERIAL TESTING LAB
47 HUDSON STREET
OSSINING, NY 10562

Cust ID: 12437

LEAK TEST CERTIFICATE

DEVICE:**Model:** 3440**Serial No:** 16813**SEALED SOURCES:**

Serial No.	Measure Date	Nuclide	GBq	mCi
50-5958	09/19/1988	Cs-137	0.296	8
47-12236	09/10/1988	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:**Sample collected on:** 12/11/2020**Sample analyzed on:** 01/06/2021 12:11:15 PM**Position:** 45**Analyzed by:** TH

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.23E+01	1.96E+01
Background measurement (cpm)	0	26
Sample measurement (cpm)	0	30
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:☒ **Less than 185 Bq (0.005 uCi)**☐ **Greater than 185 Bq (0.005 uCi)**