

(01-2021)
10 CFR 30.36(j)(1);
40.42(j)(1); 70.38(j)(1);
and 72.54(k)(5)(1)(i)



CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Library, and Information Collections Branch (T-6 A10M), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by e-mail to Infocollections.Resource@nrc.gov, and the OMB reviewer at: OMB Office of Information and Regulatory Affairs, (3150-0028), Attn: Desk Officer for the Nuclear Regulatory Commission, 725 17th Street NW, Washington, DC 20503; e-mail: olra_submission@omb.eop.gov. The NRC may not conduct or sponsor, and a person is not required to respond to, a collection of information unless the document requesting or requiring the collection displays a currently valid OMB control number.

LICENSEE NAME AND ADDRESS

GeoTesting Express, Inc.
125 Nagog Park
Acton, MA 01720

LICENSE NUMBER

20-27648-01

DOCKET NUMBER

030-35253

LICENSE EXPIRATION DATE

March 31, 2035

A. LICENSE STATUS (Check the appropriate box)

- ☐ This license has expired. ☒ This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- ☐ 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- ☐ 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner:
- ☒ a. Transfer of radioactive materials to the licensee listed below:
Commonwealth of Massachusetts / GeoTesting Express
License # 48-0214
- ☐ b. Disposal of radioactive materials:
- ☐ 1. Directly by the licensee:
- ☐ 2. By licensed disposal site:
- ☐ 3. By waste contractor:
- ☐ c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

- ☐ 1. A radiation survey was conducted by the licensee. The survey confirms:
- ☐ a. the absence of licensed radioactive materials
- ☐ b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- ☐ 2. A copy of the radiation survey results:
- ☐ a. is attached; or ☐ b. is not attached (Provide explanation); or ☐ c. was forwarded to NRC on: _____ Date _____
- ☒ 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- ☒ a. The results of the latest leak test are attached; and/or ☒ b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME Joe Tomei	TITLE Vice President & Director of Testing Services	TELEPHONE (Include Area Code) 978-893-1241	E-MAIL ADDRESS jtomei@getesting.com
-------------------	---	---	--

Mail all future correspondence regarding this license to:

Joe Tomei GeoTesting Express 125 Nagog Park Acton, MA 01720

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE

Joe Tomei Vice President &
Director of Testing Services

SIGNATURE

Joe

DATE

03/17/2021

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

INSTROTEK[®] COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.

1 Triangle Drive

P.O. Box 13944

Research Triangle Park, NC 27709

(919) 875-8371 Fax: (919) 875-8328

Mike Driscoll

Geotesting Express (Acton)

125 Nagog Park

Acton, MA (Massachusetts) 01720

Phone: 978-635-0424

LEAK TEST CERTIFICATE

InstroTek License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430

Date Swabbed: 12/17/2020

Gauge Serial Number: 70021

Source Serial Number	Reading in MicroCuries
78-9706	0.000000
77-14089	0.000202

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: _____

Customer Signature: _____

Print Date: 12/23/2020

Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.

INSTROTEK® COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.
1 Triangle Drive
P.O. Box 13944
Research Triangle Park, NC 27709
(919) 875-8371 Fax: (919) 875-8328

Mike Driscoll
Geotesting Express (Acton)
125 Nagog Park
Acton, MA (Massachusetts) 01720

Phone: 978-635-0424

LEAK TEST CERTIFICATE InstroTek License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.


Gauge Model: 3430

Date Swabbed: 12/17/2020

Gauge Serial Number: 35651

Source Serial Number	Reading in MicroCuries
78-632	0.000000
77-2708	0.000231

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: 
Customer Signature: _____

Print Date: 12/23/2020
Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.

INSTROTEK COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.
1 Triangle Drive
P.O. Box 13944
Research Triangle Park, NC 27709
(919) 875-8371 Fax: (919) 875-8328

Mike Driscoll
Geotesting Express (Acton)
125 Nagog Park
Acton, MA (Massachusetts) 01720

Phone: 978-635-0424

LEAK TEST CERTIFICATE InstroTek License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3440
Gauge Serial Number: 22946

Date Swabbed: 11/4/2020

Source Serial Number	Reading in MicroCuries
47-18105	0.000002
75-4748	0.000258

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: _____
Customer Signature: _____

Print Date: 11/11/2020
Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.

INSTROTEK COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.
1 Triangle Drive
P.O. Box 13944
Research Triangle Park, NC 27709
(919) 875-8371 Fax: (919) 875-8328

Mike Driscoll
Geotesting Express (Acton)
125 Nagog Park
Acton, MA (Massachusetts) 01720

Phone: 978-635-0424

LEAK TEST CERTIFICATE InstroTek License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3440
Gauge Serial Number: 21091

Date Swabbed: 11/20/2020

Source Serial Number	Reading in MicroCuries
47-17616	0.000000
75-2527	0.000247

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: _____
Customer Signature: _____

Print Date: 11/25/2020
Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.

INSTROTEK COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.
1 Triangle Drive
P.O. Box 13944
Research Triangle Park, NC 27709
(919) 875-8371 Fax: (919) 875-8328

Mike Driscoll
Geotesting Express (Acton)
125 Nagog Park
Acton, MA (Massachusetts) 01720

Phone: 978-635-0424

LEAK TEST CERTIFICATE InstroTek License #092-1073-1


This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3440
Gauge Serial Number: 18739

Date Swabbed: 11/20/2020

Source Serial Number	Reading in MicroCuries
47-14197	0.000000
50-8327	0.000201

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: 
Customer Signature: _____

Print Date: 11/25/2020
Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.

INSTROTEK COMPANIES



INSTROTEK • CPN • HMA • RAINHART

InstroTek Inc.
1 Triangle Drive
P.O. Box 13944
Research Triangle Park, NC 27709
(919) 875-8371 Fax: (919) 875-8328

Mike Driscoll
Geotesting Express (Acton)
125 Nagog Park
Acton, MA (Massachusetts) 01720

Phone: 978-635-0424

LEAK TEST CERTIFICATE InstroTek License #092-1073-1

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removable contamination.

Gauge Model: 3430
Gauge Serial Number: 31277

Date Swabbed: 11/4/2020

Source Serial Number	Reading in MicroCuries
47-18538	0.000000
750-5959	0.000269

Note: 0.005 MicroCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: _____
Customer Signature: _____

Print Date: 11/11/2020
Date: _____

CPN gauges are 50 mCi Am241:Be and 10 mCi Cs-137. Humboldt gauges are 40 mCi Am241:Be and 10 mCi Cs-137. InstroTek Gauge is 40 mCi Am241:Be and 10 mCi Cs-137. Troxler gauges all, except 4640, are 40 mCi Am241:Be and 8 mCi Cs-137. Troxler 4640 is 8 mCi Cs-137. Refer to the manual for other products.