

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED

Reynolds Memorial Hospital, Inc.
800 Wheeling Avenue
Glen Dale, WV 26038-1697

2. NRC/REGIONAL OFFICE

Region 1
2100 Renaissance Blvd
Suite 100
King of Prussia, PA 19406-2713

REPORT NUMBER(s) 2021001

3. DOCKET NUMBER(S)

030-03384

4. LICENSE NUMBER(S)

47-08019-01

5. DATE(S) OF INSPECTION

03/01/2021 Exit 03/29/2021

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- ☐ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☒ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

1 Non-cited violation(s) were discussed involving the following requirement(s):

10 CFR 35.61(a) requires, in part, that licensees annually calibrate the survey instruments used to show compliance with this part and 10 CFR Part 20.

Contrary to the above, in 2017, Reynolds Memorial Hospital did not annually calibrate a survey instrument used to show compliance with this part and 10 CFR Part 20. Specifically, a Ludlum 14C survey meter, serial number 214865, was calibrated on June 6, 2016, not calibrated again until August 28, 2017, and was used to perform daily surveys in Cardiology when the calibration was expired.

This is a Severity Level IV violation. (Enforcement Policy Section 6.3)

Corrective actions included calibrating the meter as well as setting up a reminder in NIMS one month before the calibration is due.

- ☐ 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with the NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11. (Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE	Michelle Snyder	<i>Michelle Snyder</i>	3/29/21
NRC INSPECTOR	Robin Elliott	<i>Robin Elliott</i>	3/29/21
BRANCH CHIEF			