

1. The licensee shall not allow eating, drinking, smoking or applying cosmetics in the licensee's areas utilizing radioactive materials.
(applies to all RAM licenses)
2. The licensee shall elute generators and process radioactive material with reagent kits in accordance with instructions furnished by the manufacturer on the label attached to or in the leaflet or brochure that accompanies the generator or reagent kit.
(applies to codes 02200, 02201, 02120)
3. The licensee shall establish written procedures for personnel performing tests to detect and quantify molybdenum-99 contamination. These procedures shall include all necessary calculations and steps to be taken if activities of molybdenum-99 in excess of the limit specified in Item 18B above are detected.
(applies to codes 02200, 02201, 02120)
 - a. For reference, parts A and B of these conditions are as follows:
 - i. 18A. The licensee shall perform a test to detect and quantify the activity of molybdenum-99 contamination in each elution of technetium-99m from a molybdenum-99/technetium-99m generator and in each extraction or separation of technetium-99m from molybdenum-99 not contained in a generator.
 - ii. 18B. The licensee shall not use technetium-99m for human use that contains more than 0.15 microcurie of molybdenum-99 per millicurie of technetium-99m. The limit for molybdenum-99 contamination represents a maximum value and molybdenum-99 contamination should be kept as low as reasonably achievable below this limit.
4. The licensee shall not transfer ownership and/or control of this license to any person or entity without providing required information, no less than thirty (30) days prior to the transfer, regarding the transfer for the Agency's review and without receiving written authorization for the transfer by the Agency.
(applies to all RAM licenses)
5. This condition is specific to the applicant's initial receipt of this license. The licensee shall notify the Radiological Health Program's Radioactive Materials Division, in writing, no later than 10 calendar days after the initial receipt of radioactive materials authorized on this license.
(applies to all RAM licenses)
6. The licensee is authorized to receive, possess, and use sealed sources of Iridium-192 where the radioactivity exceeds the maximum amount of radioactivity specified in Item 8 of this license provided:
 - a. Such possession does not exceed the quantity per source specified in Item 8 by more than 20% for Iridium-192 or 10% for Cobalt-60;
 - b. Records of the licensee show that no more than the maximum amount of radioactivity per source specified in Item 8 of the license was ordered from the

- supplier or transferor of the radioactive material; and
- c. The levels of radiation for radiographic exposure devices and storage containers do not exceed those specified in Section E. 21 of Part E "Limits on Levels of Radiation for Radiographic Exposure Devices and Storage Containers" of the Maryland Regulations COMAR 26.12.01.01 "Regulations for Control of Ionizing Radiation."
(applies to 03320 and 03310)