



September 9, 2021

Jonathan Pfingsten, Health Physicist
Commercial, Industrial, R&D and Academic Branch
Division of Radiological Safety and Security
United States Nuclear Regulatory Commission, Region 1
2100 Renaissance Blvd.
King Of Prussia, PA 19406-2713

Subject: Additional Application Information - MEAD AND HUNT, INC., REQUEST FOR ADDITIONAL INFORMATION, MAIL CONTROL NO. 627166

Dear Mr. Pfingsten:

This in reference to your letter dated August 31, 2021, requesting additional application information.

1. Regarding Item 10.6, Operating, Emergency, and Security Procedures, we would like to revise our commitment to implement the model procedures provided by the NRC in the NUREG guidance document.

We will implement and maintain the operating, emergency, and security procedures in Appendix G to NUREG-1556, Vol. 1, Rev. 2, "Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Portable Gauge Licenses." Copies of these procedures will be provided to all gauge users and will be available at each jobsite." commitment to implement the model procedures provided by the NRC in the NUREG guidance document.

2. Regarding Item 10.8, Maintenance, we would like to request to broaden the potential options for servicing and repairs to include other entities authorized to perform such work

The gauge manufacturer or other person licensed by the NRC or an Agreement State will perform nonroutine maintenance or repair operations that require detaching the source or source rod from the gauge.

If you have any questions or require additional information, please contact me.

Sincerely,
MEAD & HUNT, Inc.

A handwritten signature in blue ink that reads "James A. Bumgarner, Jr.".

James A. Bumgarner, Jr.
Vice President