



Mail Control Number: 628547
Docket Number : 3011624
License Number : 40-16775-01
Licensee Name : Prairie Lakes Healthcare System

August 26, 2021

U.S. Nuclear Regulatory Commission
Licensing Branch Region IV
1600 East Lamar Blvd.
Arlington, TX 76011-4511


To Whom It May Concern,

Please amend our Radioactive Materials License #40-16775-01 (Prairie Lakes Healthcare Systems, Inc., P.O. Box 1210, 401 9th Avenue NW; Watertown, SD 57201-6210) to include the following revision.

1. We are requesting that the duties and responsibilities of the Radiation Safety Officer be transferred to Shaun Pattrin, BA, AS. Our Administration has reviewed and approved Mr. Pattrin's credentials and experience to serve as the Radiation Safety Officer for Prairie Lakes Health Care Systems. We have enclosed the following documents for your review:
 - a. A signed Delegation of Authority that will allow Mr. Shaun Pattrin to serve as the Radiation Safety Officer at Prairie Lakes Healthcare Systems, Inc.,
 - b. A copy of Nebraska RAM License #01-133-01, verifying Mr. Pattrin has served as a Radiation Safety Officer at the Cardinal Health Cyclotron Laboratory (5840 F Street, Omaha, NE 68117) within the last seven years,
 - c. A copy of Radioactive Material License #01-59-01, verifying Mr. Pattrin as a current Authorized User for this Service License,
 - d. A copy of Nebraska Certificate of Registration #RS0050 - Services of Radiation Generating Equipment verifying Mr. Pattrin as registered personnel and includes Health Physics Consultation,
 - e. A copy of Mr. Pattrin's Certificate of Completion for PET Radiation Safety Officer Training,
2. Effective September 17, 2021, Sebastiano Anzalone, CNMT will be resigning his position as Radiation Safety Officer and will no longer be affiliated with our facility. Effective September 17, 2021, Mr. Pattrin will be assuming the duties and responsibilities of the Radiation Safety Officer, on a temporary basis until such time this amendment has been approved by your office and we have received the new radioactive materials license, naming Shaun Pattrin as the Radiation Safety Officer

Your attention in this matter is greatly appreciated.


Sincerely,



Nate Lake
Vice President
Prairie Lakes Healthcare Center

Memo To: Shaun Pattrin, BA, AS
From: Nate Lake - Vice President
Subject: Declaration of Authority

You, Shaun Pattrin, have been appointed Radiation Safety Officer and are responsible for ensuring the safe use of radiation. You are responsible for managing the radiation protection program; identifying radiation safety problems; initiating, recommending, or providing corrective actions; verifying implementation of corrective actions; stopping unsafe activities; and ensuring compliance with regulations. You are hereby delegated the authority necessary to meet those responsibilities, including prohibiting the use of byproduct material by employees who do not meet the necessary requirements and shutting down operations where justified by radiation safety. You are required to notify management of situations where staff are not cooperating and not addressing radiation safety issues. In addition, you are free to raise issues with Nuclear Regulatory Commission Regulation and Licensure at anytime. This institution also extends the authority to the Radiation Safety Officer to sign amendments in regard to this license.



Nate Lake - Vice President

9/1/21

Date

I accept the above responsibilities.



Shaun Pattrin, BA, AS

08-27-2021

Date

STATE OF NEBRASKA
DEPARTMENT OF HEALTH AND HUMAN SERVICES
DIVISION OF PUBLIC HEALTH
RADIOACTIVE MATERIALS PROGRAM

Page 1 of 3

RADIOACTIVE MATERIAL LICENSE

Pursuant to the Radiation Control Act, 1963, and 180 NAC 3 of the Nebraska regulations, 'Control of Radiation,' and in reliance on statements and representations heretofore made by the licensee designated below, a license is hereby issued authorizing such licensee to transfer, receive, possess and use the radioactive material(s) designated below; and to use such radioactive materials for the purpose(s) and at the place(s) designated below. This license is subject to all applicable rules, regulations and orders now or hereafter in effect of the Nebraska Department of Health and Human Services and to any conditions specified below.

1. Licensee Cardinal Health 414, LLC	3. License Number 01-133-01 License Type Processing or Manufacture and Distribution of radiopharmaceuticals by other than a nuclear pharmacy
2. Address 7000 Cardinal Place Dublin, OH 43017-1091	4. Amendment Number 14 License Amended In Its Entirety To Read As Follows: 5. Expiration Date January 31, 2022

6. Radioactive Material	7. Chemical And/Or Physical Form	8. Maximum Quantity License May Possess At Any One Time Under This License	9. Authorized Use
A. Fluorine-18	Any	40 curies	Production of a radiochemical and/or preparation of radioactive drug; both for transfer or distribution to authorized licenses.
<hr/>			
B. Nitrogen-13	Any	20 curies	Production of a radiochemical and/or preparation of radioactive drug; both for transfer or distribution to authorized licenses.
<hr/>			
C. Oxygen-15	Any	10 curies	Production of a radiochemical and/or preparation of radioactive drug; both for transfer or distribution to authorized licenses.
<hr/>			
D. Carbon-11	Any	10 curies	Production of a radiochemical and/or preparation of radioactive drug; both for transfer or distribution to authorized licenses.
<hr/>			
E. Any by-product material with atomic numbers 1-83 inclusive	Incidentally activated products	5 curies	Possession and storage incident to production activities.
<hr/>			
F. Any radionuclide with atomic numbers 3-83 inclusive	Sealed or solid sources manufactured under specific license issued by the U. S. Nuclear Regulatory Commission or an Agreement State,	Total not to exceed 30 millicuries, with no single source to exceed 15 millicuries.	Use for instrument calibrations and quality checks.
<hr/>			

STATE OF NEBRASKA
DEPARTMENT OF HEALTH AND HUMAN SERVICES
DIVISION OF PUBLIC HEALTH
RADIOACTIVE MATERIALS PROGRAM

Page 2 of 3

License Number: 01-133-01

Amendment Number: 14

G. Sodium-22	Sealed Source (Isotope Products Laboratories or Eckert Ziegler Model RV-22)	Total 10 millicuries, with no single source to exceed 6 millicuries.	Use for instrument calibrations and quality checks.
H. Sodium-22	Sealed Source (Isotope Products Laboratories or Eckert Ziegler Model GF-022, Type R rod)	Total 300 microcuries, with no source to exceed 220 microcuries	Use for instrument calibrations and quality checks.
I. Technetium-99m	Any unsealed form	5 curies	For instrument calibrations.

CONDITIONS

10. Licensed material shall be used by or under the supervision of: Shaun Pattrin, Robert J. Chicoine, Sean Waters, Olaf Robert Nilsson, Mehmet Husnu, Scott Lucas, Lindsay Rivera, Leonard C. Popa, Norman Medina, Jacob Kilian, Derrick Alcaide, Robert Droege, Adam Fleshner, Ken Moore, Daniel Schwede, Jeremy Guilbeau, Steve Morcos, Alyssa Rosenberg, Santiago Villegas or Charles Parraa.
11. The Radiation Safety Officer for this license is Shaun Pattrin.
12. Licensed material shall be used and/or stored only at the following location(s):
 - Cardinal Health 5840-5852 F Street Omaha, NE
13. The licensee is authorized to transport licensed material only in accordance with the provisions of 180 NAC 13.
14. The licensee is authorized to hold licensed material with a half-life of less than 120 days for decay-in-storage before disposal in ordinary trash provided:
 - A. Licensed waste to be disposed of in this manner shall be held for decay a minimum of ten (10) half-lives.
 - B. Before disposal as ordinary trash, radioactive waste shall be surveyed at the container surface with the appropriate survey meter set on its most sensitive scale and with no interposed shielding to determine that its radioactivity cannot be distinguished from background. All radiation labels shall be removed or obliterated.
 - C. A record of each disposal permitted under this Condition shall be retained for three years. The record must include the date of the disposal, the date on which the radioactive material was placed in storage, the radionuclides disposed, the survey instrument used, the background dose rate, the dose rate measured at the surface of each waste container, and the name of the individual who performed the disposal.
15. Sealed sources or detector cells containing licensed material shall not be opened or sources removed from source holders by the licensee.
16. A. Except as otherwise specifically provided by the licensee, licensed material to be distributed to licensees to administer to humans shall be procured in pre-packaged, precalibrated form from a supplier who manufactures or repackages the product under appropriate pharmaceutical controls related to assay, identity, quality, purity, sterility and nonpyrogenicity, or

STATE OF NEBRASKA
DEPARTMENT OF HEALTH AND HUMAN SERVICES
DIVISION OF PUBLIC HEALTH
RADIOACTIVE MATERIALS PROGRAM

Page 3 of 3

License Number: 01-133-01

Amendment Number: 14

- B. Licensed material prepared by the licensee may be distributed to other licensees for administration to humans, provided the product is produced under pharmaceutical controls related to assay, identity, quality, purity, sterility and nonpyrogenicity. Routine assessments shall be performed for sterility and pyrogenicity of all compounded products in use.
17. Any proposed changes in packaging, shielding, or labeling shall be submitted for review to the Radioactive Materials Program Manager, Nebraska Department of Health and Human Services, Division of Public Health, 301 Centennial Mall South, P.O. Box 95026, Lincoln, Nebraska 68509-5026.
18. The licensee shall conduct a physical inventory every six (6) months to account for all the sealed sources and/or devices received and possessed under the license. Records of the inventories shall be retained for two (2) years from the date of the inventory for inspection by the Agency. The records shall contain the model number, and serial number if one has been assigned, the identity of each source radionuclide and its nominal activity, the location of each source, and the name of the individual who performed the inventory.
19. The license does not authorize distribution pursuant to 3-014.10 or 3-014.12; to persons exempt from licensing; or to general licensees.
20. The cyclotron shall be operated by, or under the supervision and physical presence of, an authorized user listed in license condition 10.
21. Except as specifically provided otherwise by this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. Title 180 shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application with attachments dated December 16, 2016, signed by Glenn Sullivan, CRSO.
- B. Reply to deficiency letter with attachments dated January 26, 2017, signed by Glenn Sullivan, CRSO.
- C. Letter with attachment dated April 14, 2017 signed by Glenn Sullivan, CRSO showing a revised or updated floor plan for the restricted area.

Date: July 6, 2017

FOR THE NEBRASKA DEPARTMENT OF HEALTH & HUMAN SERVICES,
DIVISION OF PUBLIC HEALTH

Julia A. Schmitt
Julia A. Schmitt
Radioactive Materials Program

STATE OF NEBRASKA
DEPARTMENT OF HEALTH AND HUMAN SERVICES
DIVISION OF PUBLIC HEALTH
RADIOACTIVE MATERIALS PROGRAM

Page 1 of 2

RADIOACTIVE MATERIAL LICENSE

Pursuant to the Radiation Control Act, 1963, and 180 NAC 3 of the Nebraska regulations, 'Control of Radiation,' and in reliance on statements and representations heretofore made by the licensee designated below, a license is hereby issued authorizing such licensee to transfer, receive, possess and use the radioactive material(s) designated below; and to use such radioactive materials for the purpose(s) and at the place(s) designated below. This license is subject to all applicable rules, regulations and orders now or hereafter in effect of the Nebraska Department of Health and Human Services and to any conditions specified below.

1. Licensee	Cardinal Health 414, LLC	3. License Number	01-59-01
		License Type	Services (e.g. Leak Testing,
2. Address	5840 F Street	4. Amendment Number	41
	Omaha, NE 68111/	License Amended in Its Entirety To Read As Follows:	
		5. Expiration Date	June 30, 2022
6. Radioactive Material	7. Chemical And/Or Physical Form	8. Maximum Quantity Licensee May Possess At Any One Time Under This License	9. Authorized Use
A. Any radioactive material with atomic number 1-95, inclusive	Any	10 microcurie per isotope not to exceed 1 millicurie	For possession incident to performing leak tests as a commercial service. Any sample exceeding exempt quantities is to be returned to the original source for storage or disposal.
B. Cesium-137	Sealed Source (QSA Global, Inc. Model 77302 source rod assembly (CDC-800 Series source))	200 millicuries	Used in a QSA Global, Inc. (formerly AEA Technology-QSA Incorporated) Model 773 calibration device for calibrating instruments, including commercial calibration service.
C. Cobalt-57	Sealed Source (Eckert & Ziegler Isotope Products Model RV-057 or equivalent source from QSA Global or Bristol Myers Squibb)	10 millicuries per source. Total not to exceed 100 millicuries.	Calibrating dose testing or other radiation detection equipment.
D. Cobalt-60	Sealed Source (Bristol Myers Squibb Model NES 354 or equivalent source from Eckert & Ziegler Isotope Products)	1 millicuries per source. Total not to exceed 10 millicuries.	Calibrating dose testing or other radiation detection equipment.
E. Barium-133	Sealed Source (Bristol Myers Squibb Model NES 358 or equivalent source from Eckert & Ziegler Isotope Products)	1 millicuries per source. Total not to exceed 10 millicuries.	Calibrating dose testing or other radiation detection equipment.
F. Cesium-137	Sealed Source (Bristol Myers Squibb Model NES 356 or equivalent source from Eckert & Ziegler Isotope Products)	1 millicuries per source. Total not to exceed 10 millicuries.	Calibrating dose testing or other radiation detection equipment.
G. Samarium-153	Lexidronam liquid in syringe or vial.	300 millicuries	Calibrating dose testing or other radiation detection equipment.

STATE OF NEBRASKA
DEPARTMENT OF HEALTH AND HUMAN SERVICES
DIVISION OF PUBLIC HEALTH
RADIOACTIVE MATERIALS PROGRAM

Page 2 of 2

License Number: 01-59-01

Amendment Number: 41

H. Technetium-99m	Pertechnetate	3 curies	Calibration of nuclear medicine equipment.
I. Yttrium-90	Liquid - Salt, sealed standard	132 millicuries	Calibrating dose testing or other radiation detection equipment.
J. Radium-223	Unsealed Dichloride (Xofigo)	1 millicurie	Calibrating dose testing or other radiation detection equipment.
K. Any radioactive material with atomic number 1-83, inclusive.	Any sealed source that have been manufactured, labeled, and packaged for distribution.	100 millicuries per radionuclide. Total not to exceed 1500 millicuries.	Distribution to persons specifically licensed to receive such material.

CONDITIONS

10. Licensed Material shall only be used by, or under the supervision of Frank Rutar, Sebastiano Anzalone, Kendra Pack, Shaun Pattrin, Krista Kuhlman, Janet Egenberger, and Upendra Adhikari.
11. The Radiation Safety Officer for this license is Sebastiano Anzalone.
12. Licensed material shall be used and/or stored only at the following location(s) and temporary jobsites:
- Cardinal Health 414, LLC 5840 F Street Omaha, NE
13. The licensee is authorized to transport licensed material only in accordance with the provisions of 180 NAC 13.
14. Sealed sources containing licensed material shall not be opened by the licensee.
15. Except as specifically provided otherwise by this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. Title 180 shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
 - A. Renewal application with attachments dated May 21, 2015, signed by Duane Kanne, Director.

Date: 4-12-21

FOR THE NEBRASKA DEPARTMENT OF HEALTH & HUMAN SERVICES,
DIVISION OF PUBLIC HEALTH


Becky Wisell
Radioactive Materials Program

NEBRASKA

Good Life. Great Mission.

DEPT. OF HEALTH AND HUMAN SERVICES

Registration Number: RS0050

Expiration Date: March 31, 2022

CERTIFICATE OF REGISTRATION

Services of Radiation Generating Equipment

Registration #: RS0050

Cardinal Health 414, LLC
Medical/Health Physics Dept.
Omaha NE 68117

Phone: 402-734-8045 #6

Fax: 402-553-3033

E-mail: sebastiano.anzalone@cardinalhealth.com

Responsible Contact:

Sebastiano Anzalone

Contact Phone: 402-734-8045 #6

Service Types

A. Installation/Service:

- ☐ A1 Installation/Assembly (including initial Electronic Calibration) Radiation Generating Equipment
- ☐ A2 Service/repair of Radiation Generating Equipment
- ☒ A3 Measurement of Radiation Generating Equipment Output

B. Calibration:

- ☒ B1 Calibration of Diagnostic Radiation Generating Equipment
- ☐ B2 Calibration of CTs
- ☐ B3 Calibration of Therapeutic Radiation Generating Equipment
- ☐ B4 Calibration of Non-Medical Radiation Generating Equipment

C. Reviews:

- ☒ C1 Area surveys and shielding reviews of Diagnostic Radiation Generating Facilities
- ☒ C2 CT Shielding Reviews
- ☒ C3 Therapeutic Reviews
- ☒ C4 Non-Medical area survey and shielding reviews of Radiation Generating Facilities
- ☐ D. Demonstration which includes energizing the radiation generating equipment
- ☐ E. Sales
- ☒ F. Other: - Consultations Industrial & On-Site Physician Training Course

Registration does not imply approval or disapproval of service.

As a registrant you are required to notify this Department in writing within thirty (30) days of any changes which would render the information contained in the application of registration no longer accurate per 180 NAC 2-008.01.

CERTIFICATE OF REGISTRATION

Services of Radiation Generating Equipment

Registered Personnel

Personnel qualified to perform each service listed above. Document training in accordance with 180 NAC 15, Nebraska Regulations for Control of Radiation-Ionizing. Notify this Department of additional names if necessary.

Name: **Upendra Adhikari**

Approved
Service
Types

A A1 ☐ A2 ☐ A3 ☒

B B1 ☒ B2 ☐ B3 ☐ B4 ☐

C C1 ☒ C2 ☐ C3 ☐ C4 ☐

D ☐ E ☒ F ☒

☐ Check to remove Qualified Individual

Name: **Sebastiano Anzalone**

Approved
Service
Types

A A1 ☐ A2 ☐ A3 ☒

B B1 ☒ B2 ☐ B3 ☐ B4 ☐

C C1 ☒ C2 ☐ C3 ☐ C4 ☐

D ☐ E ☒ F ☒

☐ Check to remove Qualified Individual

Name: **Janet Egenberger**

Approved
Service
Types

A A1 ☐ A2 ☐ A3 ☒

B B1 ☒ B2 ☐ B3 ☐ B4 ☐

C C1 ☒ C2 ☐ C3 ☐ C4 ☐

D ☐ E ☒ F ☒

☐ Check to remove Qualified Individual

Name: **Krista Kuhlman**

Approved
Service
Types

A A1 ☐ A2 ☐ A3 ☒

B B1 ☒ B2 ☐ B3 ☐ B4 ☐

C C1 ☒ C2 ☐ C3 ☐ C4 ☐

D ☐ E ☐ F ☒

☐ Check to remove Qualified Individual

Name: **Kendra Pack**

Approved
Service
Types

A A1 ☐ A2 ☐ A3 ☒

B B1 ☒ B2 ☐ B3 ☐ B4 ☐

C C1 ☒ C2 ☐ C3 ☐ C4 ☐

D ☐ E ☐ F ☒

☐ Check to remove Qualified Individual

Registration does not imply approval or disapproval of service.

As a registrant you are required to notify this Department in writing within thirty (30) days of any changes which would render the information contained in the application of registration no longer accurate per 180 NAC 2-008.01.

~~NEBRASKA~~

Good Life. Great Mission.

DEPT. OF HEALTH AND HUMAN SERVICES

Registration Number: RS0050

Expiration Date: March 31, 2022

CERTIFICATE OF REGISTRATION
Services of Radiation Generating Equipment

Name: **Shaun** **Patrin**

Approved
Service

A A1 ☐ A2 ☐ A3 ☒

B B1 ☒ B2 ☐ B3 ☐ B4 ☐

Types: C C1 ☒ C2 ☐ C3 ☐ C4 ☐

D ☐ E ☐ F ☒

☐ Check to remove Qualified Individual

Name: **Frank** **Rutar**

Approved
Service

A A1 ☐ A2 ☐ A3 ☒

B B1 ☒ B2 ☐ B3 ☐ B4 ☐

Types: C C1 ☒ C2 ☒ C3 ☒ C4 ☒

D ☐ E ☐ F ☒

☐ Check to remove Qualified Individual

Registration does not imply approval or disapproval of service.

As a registrant you are required to notify this Department in writing within thirty (30) days of any changes which would render the information contained in the application of registration no longer accurate per 180 NAC 2-008.01.

Certificate of Completion

Quality & Regulatory Department

certifies that

Shaun Pattrin

successfully completed the

**PET Radiation Safety Officer
Training Course**

on October 29, 2010 in Dublin, OH



CardinalHealth



Director, Health Physics, Quality & Regulatory

From: [Tomczak, Tammy](#)
To: [Hill, Carol](#)
Subject: FW: RSO Responsibility transfer
Date: Thursday, September 2, 2021 9:11:48 AM
Attachments: [DOC090221.pdf](#)
Importance: High

Good morning, Carol,

It looks like one of our reviewers got this amendment request, which is for Region IV. I have not done anything with it in WBL, because I didn't want to mess up your process.

Please let me know if you need anything else from me.

Thank!!
Tammy

From: Cender, Laura <Laura.Cender@nrc.gov>
Sent: Thursday, September 02, 2021 9:08 AM
To: Tomczak, Tammy <Tammy.Tomczak@nrc.gov>
Cc: Pavon, Sandy <Sandy.Pavon@nrc.gov>
Subject: FW: RSO Responsibility transfer

Hello Tammy,

Please process the attached file in as a new license amendment request for Prairie Lakes Healthcare System. This is an RIV licensee and will need to be forwarded to them for case assignment.

License No. 40-16775-01
Docket No. 030-11624

Please let me know if you have any questions.

Thank you,
Laura

From: Bryan Beckedahl <Bryan.Beckedahl@prairielakes.com>
Sent: Thursday, September 02, 2021 9:01 AM
To: Cender, Laura <Laura.Cender@nrc.gov>
Cc: Anzalone, Sebastiano <Sebastiano.Anzalone@cardinalhealth.com>
Subject: [External_Sender] RSO Responsibility transfer

Greetings,
Attached you will find the required documentation for transferring the responsibilities of the Radiation Safety Officer.
Let me know if you have any questions and thank you for your time.

Cordially,

Bryan Beckedahl
Director of Radiology
Prairie Lakes Healthcare System
401 9th Ave. NW Watertown, SD 57201
605-882-7986
Bryan.Beckedahl@prairielakes.com





ACKNOWLEDGEMENT - RECEIPT OF CORRESPONDENCE

Name and Address of Applicant and/or Licensee

Jenna Hulscher, CRA
Director of Radiology
Prairie Lakes Healthcare System
401 9th Ave NW
P.O. Box 1210
Watertown, SD 57201-6210

Date

09/15/2021

License Number(s)

40-16775-01

Mail Control Number(s)

628547

Licensing and/or Technical Reviewer or Branch

C. Hill

This is to acknowledge receipt of your: ☒ Letter and/or ☐ Application Dated: 08/26/2021

The initial processing, which included an administrative review, has been performed.

☒ Amendment ☐ Termination ☐ New License ☐ Renewal

☐ There were no administrative omissions identified during our initial review.

☐ This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

☐ Your application for a new NRC license did not include your taxpayer identification number. Please complete and submit NRC Form 531, Request for Taxpayer Identification Number, located at the following link: <http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>
Follow the instructions on the form for submission.

☐ The following administrative omissions have been identified:

Your application has been assigned the above listed MAIL CONTROL NUMBER. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region IV
U. S. Nuclear Regulatory Commission
DNMS/NMSB - B
1600 E. Lamar Boulevard
Arlington, TX 76011-4511
(817) 200-1103 or (817) 200-1140

BETWEEN:

Accounts Receivable/Payable
and
Regional Licensing Branches

[FOR ARPB USE]
INFORMATION FROM WBL

Program Code: 02120
Status Code: Pending Amendment
Fee Category: 7C
Exp. Date: 11/30/2023
Fee Comments: CODE 23
Decom Fin Assur Req: N

License Fee Worksheet - License Fee Transmittal

A. REGION

1. APPLICATION ATTACHED

Applicant/Licensee: Prairie Lakes Healthcare System
Received Date: 09/02/2021
Docket Number: 3011624
Mail Control Number: 628547
License Number: 40-16775-01
Action Type: Amendment

2. FEE ATTACHED

Amount: N/A

Check No.: N/A

3. COMMENTS

Signed: Carol L. Hill

Date: 09/15/2021

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / /)

1. Fee Category and Amount: _____

2. Correct Fee Paid. Application may be processed for:

Amendment: _____

Renewal: _____

License: _____

3. OTHER _____

Signed: _____

Date: _____

Agency: NRC

WBL WORKSHEET

DOCKET NUMBER: 3011624	LICENSE NUMBER: 40-16775-01	STATUS: Pending Amendment	
MAIL CONTROL NUMBER: 628547	RECEIPT DATE: 09/02/2021	ACTION TYPE: Amendment	
DUE DATE: 12/01/2021	INST. CODE: 16775	LICENSE REGION: Region 4	
LICENSE TYPE: 30	ENTITY TYPE: C	LICENSE GROUP: Medical	
ISSUE DATE:	ORIGINAL DATE: 11/07/1988	EXPIRATION DATE: 11/30/2023	
DECOMMISSIONING CATEGORY: Group 1		LAST ISSUE DATE:	
LICENSEE NAME: Prairie Lakes Healthcare System		DECOM FIN ASSUR REQD: N SUBM: N	
MAILING ADDRESS LINE1: 401 9th Avenue Northwest		CONT PLAN REQD: N APPRV: N	
MAILING ADDRESS LINE 2: P.O. Box 1210			
CITY: Watertown	STATE: SD	ZIP: 57201-6210	
CONTACT PERSON: PREFIX:	FIRST NAME: Jenna	MIDDLE INITIAL:	
LAST NAME: Hulscher	SUFFIX: CRA		
JOB TITLE: Director of Radiology	PHONE: 605-882-7986	FAX: 605-882-7070	EMAIL: jenna.hulscher@prairie
BILLING ADDRESS LINE 1:			
BILLING ADDRESS LINE 2:			
CITY:	STATE: South Dakota	ZIP:	
BILLING CONTACT PERSON: FIRST NAME:	MIDDLE INITIAL:	LAST NAME:	
PHONE:	EMAIL:	FAX:	
PRIMARY PGM CODE: 02120	SECONDARY PGM CODE:		
INSPECTION REGION: Region 4	PRIORITY: 3		
RSO: PREFIX:	FIRST NAME: Sebastiano	MIDDLE INITIAL:	LAST NAME Anzalone
SUFFIX: BS, CNMT, AART	RSO JOB TITLE:		
RSO PHONE: 605-882-7986	RSO FAX:	RSO EMAIL: sebastiano.anzalone@prairielakes.com	
STATES WHERE USE IS AUTHORIZED: 1		0- ALL LISTED STATES 1- SAME AS STATE IN ADDRESS 2- ALL STATES 3- NON-AGREEMENT-STATES	
AUTHORIZED STATES (USE ONLY IF ABOVE IS ZERO):			