



A Hartford HealthCare Partner

80 Seymour Street Hartford, CT 06106
24 September 2021

To: Robin Elliott
Senior Health Physicist
U.S. Nuclear Regulatory Commission
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406

From: Bette W. Blankenship M.S., D.A.B.R.
Medical Physicist and Radiation Safety Officer
Hartford Hospital
Department of Radiation Safety

RAM License No. 06-0025304

Dear Ms. Elliott;

In response to your email dated September 21st 2021 requesting additional source manufacturer and source specifics to add Cs-131 to the hospital prostate implant program:

Sealed source regulatory Information
Isoray Cs-131 Model CS-1
510(K): K030162
Regulation 21 CFR 892.5730
Radionuclide brachytherapy source
Regulatory Class II
Product code: 90 KXX
Approval received: March 28, 2003

Sealed Source and Device Registry Number: WA-1220-S-101-S
Approved September 17, 2004

Possession request: 2 Ci total activity, 10 mCi per seed maximum

Please contact me with any questions.

Thank you,

Bette W. Blankenship M.S., DABR, FAAPM
Medical Physicist and Radiation Safety Officer
Hartford Hospital
80 Seymour Street
Hartford, CT 06106
bette.blankenship@hhchealth.org