

**From:** [Tran, Frank](#)  
**To:** [ebell2@ecommunity.com](mailto:ebell2@ecommunity.com)  
**Subject:** Request for additional information for NRC License No. 13-06009-01  
**Date:** Friday, November 05, 2021 3:54:00 PM

---

Dear Ms. Bell:

This refers to the license amendment request dated October 22, 2021 for Community Health Network, Inc., NRC License No. 13-06009-01. We reviewed the application and will need the following.

1. It appears the licensing guidance cited in Item g of your application was out dated. The current licensing guidance for your license is NUREG-1556, Volume 9, **Revision 3**. Please correct and resubmit the commitment stated in Item g.
2. The regulation cited in Item i of your application was incorrect. Based on NUREG-1556, Volume 9, Rev.3, the statement should be "We have developed and will implement and maintain written procedures for safe use of unsealed byproduct material that meet the requirements of 10 CFR 20.1101 and 10 CFR **20.1201**."
3. Please confirm if PET isotopes will be used at your facilities located at 1400 N. Ritter Ave., Indianapolis, Indiana, 8075 N. Shadeland Ave., Indianapolis, Indiana, and 1402 E. County Line Rd., Indianapolis, Indiana. If PET isotopes will be used, please provide information as discussed in Item 9 of Appendix C to NUREG-1556, Vol. 9, Rev.3 (shielding evaluation including information about the type, thickness, and density of any necessary shielding to enable independent verification of shielding calculations, and a description of any portable shields used. The calculations should include the workload assumptions used. A description of surrounding areas, including the occupancy factors, and indicate whether the areas are restricted or unrestricted (for calculations of the maximum exposure in any given hour, an occupancy factor should not be used)).
4. In order to transfer all three locations in Item 3 above from NRC License No. 13-24834-01 to your license, we will need to terminate NRC License No. 13-24834-01. Please determine if this license will be terminated or not and submit a NRC Form 314 (sign by the licensee of NRC License No. 13-24834-01) as necessary. A copy of the form is attached for your reference.
5. If NRC License No. 13-24834-01 will be terminated, please confirm that all records required to maintain by the NRC regulations under NRC License No. 13-24834-01 will be transferred to NRC License No. 13-06009-01.

To continue the review of your application, we request that you submit your response under a dated and signed cover letter by December 6, 2021. In the cover letter, please refer the license number, docket number and Mail Control No. 629011. We will assume that you do not wish to further pursue this licensing action if we do not receive a reply within the specified timeframe noted above.

If you have questions, require additional time to respond, or require clarification on any of the information stated above, please contact me at 630-829-9623 or reply to this email.

In accordance with Title 10 of the *Code of Federal Regulations* (10 CFR) 2.390 of the NRC's "Rules of Practice," a copy of this correspondence will be made available electronically for public inspection in the NRC Public Document Room or from the NRC's Agencywide Documents Access and Management System (ADAMS), accessible from the NRC Web site at <https://www.nrc.gov/reading-rm/adams.html>.

Best regards,

*Frank Tran*

Health Physicist/License Reviewer  
NRC Region III/Division of Nuclear Materials Safety  
Phone: 630-829-9623  
Fax: 630-515-1078  
Email: [Frank.Tran@nrc.gov](mailto:Frank.Tran@nrc.gov)

