



Materials Inspection Record

1. Licensee Name: Academic Cardiology Associates, P.C.		2. Docket Number(s): 030-35278		3. License Number(s): 21-32228-02	
4. Report Number(s): 2022-001			5. Date(s) of Inspection: January 4, 2022		
6. Inspector(s): D. A. Piskura		7. Program Code(s): 02201		8. Priority: 5	9. Inspection Guidance Used: 87130
10. Licensee Contact Name(s): Lara Luna, RSO		11. Licensee E-mail Address: lluna@mpcphysics.com		12. Licensee Telephone Number(s): 734-662-3197	
13. Inspection Type: <input checked="" type="checkbox"/> Routine <input type="checkbox"/> Announced <input type="checkbox"/> Non-Routine <input checked="" type="checkbox"/> Unannounced		14. Locations Inspected: <input checked="" type="checkbox"/> Main Office <input type="checkbox"/> Field Office <input type="checkbox"/> Temporary Job Site <input type="checkbox"/> Remote		15. Next Inspection Date (MM/DD/YYYY): 01/04/2027 <input checked="" type="checkbox"/> Normal <input type="checkbox"/> Extended <input type="checkbox"/> Reduced <input type="checkbox"/> No change	

16. Scope and Observations:

This licensee was a private cardiology out-patient clinic, authorized to use licensed material permitted by Sections 35.100 and 35.200. The nuclear medicine department was staffed with one technologist who performed approximately 100 diagnostic procedures monthly. The licensee received unit doses from a licensed radiopharmacy. The clinic administered studies on four days each week on Mondays, Tuesdays, Wednesdays, and Fridays. The clinic retained the services of a consulting physicist who also served as the RSO. The consultant/RSO audited the radiation safety program on a quarterly basis (last on 10/4/2021 with no findings).

This inspection consisted of interviews with licensee personnel, a review of selected records, a tour of the nuclear medicine department, and independent measurements. At the time of this inspection, all patient procedures had been completed and the inspector was unable to observe the administration of testing procedures. The inspector observed licensee personnel demonstrate dose calibrator QA tests, package receipt and surveys, and area radiation surveys.

No violations of NRC requirements were identified during this inspection.

MK
1-13-2022